

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041420\
 Data File : PS009757.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2020 22:37
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 05:22:02 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.238	6.928	441.7E6	375.4E6	798.547	683.521
Target Compounds							
1) T	Dalapon	3.046	2.629	484.6E6	468.1E6	621.437	594.561
2) T	3,5-DICHL...	6.448	6.075	552.1E6	516.0E6	731.478	634.400
3) T	4-Nitroph...	7.067	6.531	195.4E6	209.7E6	818.357	658.944
5) T	DICAMBA	7.431	7.098	1553.8E6	1419.2E6	730.115	618.239
6) T	MCPD	7.601	7.196	123.0E6	117.4E6	49.989	64.573 #
7) T	MCPA	7.751	7.412	174.9E6	169.2E6	57.411	62.697
8) T	DICHLORPROP	8.124	7.725	377.0E6	358.7E6	730.920	646.520
9) T	2,4-D	8.353	8.018	405.3E6	433.4E6	768.829	659.310
10) T	Pentachlo...	8.667	8.452	5120.0E6	4373.0E6	728.185	576.149
11) T	2,4,5-TP ...	9.215	8.812	2063.3E6	1910.1E6	752.693	640.719
12) T	2,4,5-T	9.499	9.191	2008.5E6	1933.2E6	789.551	654.824
13) T	2,4-DB	10.052	9.698	281.0E6	261.2E6	812.390	687.957
14) T	DINOSEB	11.218	10.033	1510.2E6	1412.3E6	738.776	642.991
15) T	Picloram	11.030	11.012	2021.1E6	2128.6E6	665.933	522.711
16) T	DCPA	11.482	10.963	2570.7E6	2465.7E6	741.627	636.635

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041420\
Data File : PS009757.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2020 22:37
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 05:22:02 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

