

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041422\
 Data File : PS018989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2022 19:47
 Operator : AJ\MA
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 15 03:23:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041422.M
 Quant Title : 8080.M
 QLast Update : Fri Apr 15 02:14:56 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.230	6.806	1105.1E6	344.7E6	312.654	453.882 #
Target Compounds							
9) T	2,4-D	8.417f	0.000	24187633	0	6.081	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041422\
Data File : PS018989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2022 19:47
Operator : AJ\MA
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 15 03:23:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041422.M
Quant Title : 8080.M
QLast Update : Fri Apr 15 02:14:56 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.230 min
Delta R.T.: 0.001 min
Response: 1105127544
Conc: 312.65 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.806 min
Delta R.T.: 0.000 min
Response: 344745001
Conc: 453.88 ng/ml



#9 2,4-D

R.T.: 8.417 min
Delta R.T.: 0.063 min
Response: 24187633
Conc: 6.08 ng/ml



#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.891 min
Response: 0
Conc: N.D.