

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041621\
 Data File : PS014677.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Apr 2021 19:19
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 00:49:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.157	6.884	1159.6E6	358.2E6	762.309	769.497
Target Compounds							
1) T	Dalapon	3.007	2.603	1079.1E6	384.7E6	598.430	591.127
2) T	3,5-DICHL...	6.383	6.046	1474.3E6	519.2E6	711.550	708.080
3) T	4-Nitroph...	6.991	6.499	559.3E6	202.7E6	688.206	718.032
5) T	DICAMBA	7.346	7.050	4196.5E6	1458.0E6	721.252	720.866
6) T	MCPD	7.513	7.146	318.7E6	122.2E6	73.190	72.582
7) T	MCPA	7.662	7.361	446.4E6	171.7E6	71.740	70.722
8) T	DICHLORPROP	8.031	7.675	1013.7E6	364.6E6	724.420	722.305
9) T	2,4-D	8.260	7.971	1129.0E6	427.3E6	710.081	768.447
10) T	Pentachlo...	8.566	8.394	15159.7E6	4922.5E6	721.314	732.241
11) T	2,4,5-TP ...	9.111	8.755	5946.5E6	2007.5E6	726.387	734.906
12) T	2,4,5-T	9.395	9.135	5948.2E6	1959.3E6	723.087	749.089
13) T	2,4-DB	9.943	9.641	1055.0E6	257.8E6	693.282	743.180
14) T	DINOSEB	11.096	9.971	4512.2E6	1402.6E6	700.261	728.365
15) T	Picloram	10.920	10.955	7075.3E6	2389.0E6	632.380	637.077
16) T	DCPA	11.361	10.898	8893.3E6	3021.8E6	721.739	744.158

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041621\
Data File : PS014677.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Apr 2021 19:19
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 00:49:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

