

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041720\
 Data File : PS009852.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2020 22:54
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 08:37:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 12 07:13:24 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	Dalapon	779.834	693.970 E3	11.0	86	0.00
2 T	3,5-DICHLORO BENZOIC ACID	754.798	815.544 E3	-8.0	104	0.00
3 T	4-Nitrophenol	238.713	288.310 E3	-20.8#	116	0.00
4 S	2,4-DCAA	553.170	604.942 E3	-9.4	106	0.00
5 T	DICAMBA	2.128	2.285 E6	-7.4	103	0.00
6 T	MCP	2.461	1.824 E6	25.9#	73	0.00
7 T	MCPA	3.046	2.613 E6	14.2	82	0.00
8 T	DICHLORPROP	515.743	552.288 E3	-7.1	103	0.00
9 T	2,4-D	527.225	584.902 E3	-10.9	107	0.00
10 T	Pentachlorophenol	7.031	7.399 E6	-5.2	102	0.00
11 T	2,4,5-TP (SILVEX)	2.741	3.016 E6	-10.0	106	0.00
12 T	2,4,5-T	2.544	2.862 E6	-12.5	108	0.00
13 T	2,4-DB	345.955	395.490 E3	-14.3	110	0.00
14 T	DINOSEB	2.044	2.205 E6	-7.9	104	0.00
15 T	Picloram	3.035	2.918 E6	3.9	90	0.00
16 T	DCPA	3.466	3.704 E6	-6.9	103	0.00

Signal #2

1 T	Dalapon	787.363	666.333 E3	15.4#	82	0.00
2 T	3,5-DICHLORO BENZOIC ACID	813.307	754.245 E3	7.3	90	0.00
3 T	4-Nitrophenol	318.198	310.039 E3	2.6	94	0.00
4 S	2,4-DCAA	549.198	510.861 E3	7.0	91	0.00
5 T	DICAMBA	2.296	2.057 E6	10.4	86	0.00
6 T	MCP	1.819	1.697 E6	6.7	89	0.00
7 T	MCPA	2.699	2.481 E6	8.1	88	0.00
8 T	DICHLORPROP	554.822	516.781 E3	6.9	91	0.00
9 T	2,4-D	657.361	621.649 E3	5.4	92	-0.01
10 T	Pentachlorophenol	7.590	6.517 E6	14.1	84	-0.01
11 T	2,4,5-TP (SILVEX)	2.981	2.742 E6	8.0	89	-0.01
12 T	2,4,5-T	2.952	2.748 E6	6.9	90	-0.01
13 T	2,4-DB	379.616	365.965 E3	3.6	94	-0.01
14 T	DINOSEB	2.196	2.020 E6	8.0	89	-0.01
15 T	Picloram	4.072	3.097 E6	23.9#	72	-0.01
16 T	DCPA	3.873	3.501 E6	9.6	88	-0.01

Evaluate Continuing Calibration Report - Not Found

Signal #2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041720\
Data File : PS009852.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2020 22:54
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 08:37:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS031120.M
Quant Title : 8080.M
QLast Update : Thu Mar 12 07:13:24 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0