

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041720\
 Data File : PS009838.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2020 17:33
 Operator : DD\AJ
 Sample : L2281-14
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 01:54:53 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.297	6.948	6942032	17042905	11.497	34.413 #
Target Compounds							
1) T	Dalapon	0.000	2.588	0	5104907	N.D.	7.131
2) T	3,5-DICHL...	0.000	6.105	0	96187675	N.D.	129.475
3) T	4-Nitroph...	7.126	6.516	6662498	1451156	23.949	4.779 #
5) T	DICAMBA	7.420	7.061	6086929	20562324	2.702	10.338 #
6) T	MCPD	7.653	7.231	13625401	1869791	7.631	1.136 #
7) T	MCPA	7.734	7.410	34755232	3476362	13.483	1.457 #
8) T	DICHLORPROP	0.000	7.745	0	14019143	N.D.	27.854
9) T	2,4-D	8.422	7.998	11615635	2221229	20.230	3.707 #
10) T	Pentachlo...	0.000	8.442	0	1986635	N.D.	<MDL
11) T	2,4,5-TP ...	0.000	8.778	0	80115492	N.D.	30.634
12) T	2,4,5-T	9.445	9.261	50878784	8870778	18.350	3.399 #
13) T	2,4-DB	0.000	9.637	0	8280690	N.D.	23.354
14) T	DINOSEB	0.000	10.077	0	97288843	N.D.	49.404
15) T	Picloram	11.007	11.034	89083282	4591057	28.754	1.350 #
16) T	DCPA	0.000	10.970	0	2132741	N.D.	<MDL


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041720\
Data File : PS009838.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2020 17:33
Operator : DD\AJ
Sample : L2281-14
Misc :
ALS Vial : 20 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 01:54:53 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 3.040 min
Response: 0
Conc: N.D.




#1 Dalapon

R.T.: 2.588 min
Delta R.T.: -0.033 min
Response: 5104907
Conc: 7.13 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.448 min
Response: 0
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.105 min
Delta R.T.: 0.031 min
Response: 96187675
Conc: 129.47 ng/ml




#3 4-Nitrophenol

R.T.: 7.126 min
Delta R.T.: 0.058 min
Response: 6662498
Conc: 23.95 ng/ml




#3 4-Nitrophenol

R.T.: 6.516 min
Delta R.T.: -0.014 min
Response: 1451156
Conc: 4.78 ng/ml




#4 2,4-DCAA

R.T.: 7.297 min
Delta R.T.: 0.058 min
Response: 6942032
Conc: 11.50 ng/ml




#4 2,4-DCAA

R.T.: 6.948 min
Delta R.T.: 0.020 min
Response: 17042905
Conc: 34.41 ng/ml




#5 DICAMBA

R.T.: 7.420 min
Delta R.T.: -0.012 min
Response: 6086929
Conc: 2.70 ng/ml




#5 DICAMBA

R.T.: 7.061 min
Delta R.T.: -0.037 min
Response: 20562324
Conc: 10.34 ng/ml




#6 MCP

R.T.: 7.653 min
Delta R.T.: 0.052 min
Response: 13625401
Conc: 7.63 ug/ml




#6 MCP

R.T.: 7.231 min
Delta R.T.: 0.036 min
Response: 1869791
Conc: 1.14 ug/ml




#7 MCPA

R.T.: 7.734 min
Delta R.T.: -0.017 min
Response: 34755232
Conc: 13.48 ug/ml




#7 MCPA

R.T.: 7.410 min
Delta R.T.: -0.001 min
Response: 3476362
Conc: 1.46 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.125 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 7.745 min
Delta R.T.: 0.020 min
Response: 14019143
Conc: 27.85 ng/ml




#9 2,4-D

R.T.: 8.422 min
Delta R.T.: 0.068 min
Response: 11615635
Conc: 20.23 ng/ml




#9 2,4-D

R.T.: 7.998 min
Delta R.T.: -0.021 min
Response: 2221229
Conc: 3.71 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T. : 9.216 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 8.778 min
Delta R.T.: -0.034 min
Response: 80115492
Conc: 30.63 ng/ml




#12 2,4,5-T

R.T.: 9.445 min
Delta R.T.: -0.056 min
Response: 50878784
Conc: 18.35 ng/ml




#12 2,4,5-T

R.T.: 9.261 min
Delta R.T.: 0.070 min
Response: 8870778
Conc: 3.40 ng/ml




#13 2,4-DB

R.T.: 0.000 min
Exp R.T.: 10.053 min
Response: 0
Conc: N.D.




#13 2,4-DB

R.T.: 9.637 min
Delta R.T.: -0.062 min
Response: 8280690
Conc: 23.35 ng/ml




#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 11.219 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 10.077 min
Delta R.T.: 0.043 min
Response: 97288843
Conc: 49.40 ng/ml



#15 Picloram

R.T.: 11.007 min
Delta R.T.: -0.025 min
Response: 89083282
Conc: 28.75 ng/ml



#15 Picloram

R.T.: 11.034 min
Delta R.T.: 0.021 min
Response: 4591057
Conc: 1.35 ng/ml