

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041720\
 Data File : PS009847.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2020 20:59
 Operator : DD\AJ
 Sample : L2284-08
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 01:57:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
4) S	2,4-DCAA	7.238	6.927	359.1E6	340.8E6	594.783	688.184
Target Compounds							
1) T	Dalapon	3.004	2.584	19640435	77405517	25.924	108.125 #
2) T	3,5-DICHL...	6.488	6.105	268307	75872440	<MDL	102.129 #
3) T	4-Nitroph...	7.127	6.558	26282600	2329471	94.475	7.671 #
5) T	DICAMBA	7.423	7.094	7304692	2945170	3.243	1.481 #
6) T	MCPD	0.000	7.226	0	1983255	N.D.	1.205
7) T	MCPA	7.708	7.404	31651185	2934970	12.279	1.230 #
8) T	DICHLORPROP	0.000	7.745	0	19952391	N.D.	39.642
9) T	2,4-D	8.422	8.037	17020404	3490553	29.643	5.825 #
10) T	Pentachlo...	0.000	8.446	0	6149138	N.D.	<MDL
11) T	2,4,5-TP ...	0.000	8.783	0	86496365	N.D.	33.074
12) T	2,4,5-T	9.444	9.190	37103272	3645434	13.382	1.397 #
13) T	2,4-DB	0.000	9.702	0	1737651	N.D.	4.901
14) T	DINOSEB	0.000	10.076	0	115.9E6	N.D.	58.848
15) T	Picloram	11.007	11.012	62407647	9661531	20.144	2.841 #
16) T	DCPA	0.000	10.961	0	5195911	N.D.	1.550


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041720\
Data File : PS009847.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2020 20:59
Operator : DD\AJ
Sample : L2284-08
Misc :
ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 01:57:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon

R.T.: 3.004 min
Delta R.T.: -0.036 min
Response: 19640435
Conc: 25.92 ng/ml




#1 Dalapon

R.T.: 2.584 min
Delta R.T.: -0.037 min
Response: 77405517
Conc: 108.12 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.488 min
Delta R.T.: 0.039 min
Response: 268307
Conc: N.D.




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.105 min
Delta R.T.: 0.030 min
Response: 75872440
Conc: 102.13 ng/ml




#3 4-Nitrophenol

R.T.: 7.127 min
Delta R.T.: 0.059 min
Response: 26282600
Conc: 94.48 ng/ml




#3 4-Nitrophenol

R.T.: 6.558 min
Delta R.T.: 0.028 min
Response: 2329471
Conc: 7.67 ng/ml




#4 2,4-DCAA

R.T.: 7.238 min
Delta R.T.: 0.000 min
Response: 359139125
Conc: 594.78 ng/ml




#4 2,4-DCAA

R.T.: 6.927 min
Delta R.T.: -0.001 min
Response: 340823352
Conc: 688.18 ng/ml




#5 DICAMBA

R.T.: 7.423 min
Delta R.T.: -0.009 min
Response: 7304692
Conc: 3.24 ng/ml




#5 DICAMBA

R.T.: 7.094 min
Delta R.T.: -0.004 min
Response: 2945170
Conc: 1.48 ng/ml




#6 MCP

R.T.: 0.000 min
Exp R.T.: 7.602 min
Response: 0
Conc: N.D.




#6 MCP

R.T.: 7.226 min
Delta R.T.: 0.031 min
Response: 1983255
Conc: 1.20 ug/ml




#7 MCPA

R.T.: 7.708 min
Delta R.T.: -0.043 min
Response: 31651185
Conc: 12.28 ug/ml




#7 MCPA

R.T.: 7.404 min
Delta R.T.: -0.008 min
Response: 2934970
Conc: 1.23 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.125 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 7.745 min
Delta R.T.: 0.020 min
Response: 19952391
Conc: 39.64 ng/ml




#9 2,4-D

R.T.: 8.422 min
Delta R.T.: 0.068 min
Response: 17020404
Conc: 29.64 ng/ml




#9 2,4-D

R.T.: 8.037 min
Delta R.T.: 0.018 min
Response: 3490553
Conc: 5.83 ng/ml




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T.: 9.216 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 8.783 min
Delta R.T.: -0.029 min
Response: 86496365
Conc: 33.07 ng/ml




#12 2,4,5-T

R.T.: 9.444 min
Delta R.T.: -0.057 min
Response: 37103272
Conc: 13.38 ng/ml




#12 2,4,5-T

R.T.: 9.190 min
Delta R.T.: -0.002 min
Response: 3645434
Conc: 1.40 ng/ml




#13 2,4-DB

R.T.: 0.000 min
Exp R.T.: 10.053 min
Response: 0
Conc: N.D.




#13 2,4-DB

R.T.: 9.702 min
Delta R.T.: 0.003 min
Response: 1737651
Conc: 4.90 ng/ml




#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 11.219 min
Response: 0
Conc: N.D.




#14 DINOSEB

R.T.: 10.076 min
Delta R.T.: 0.042 min
Response: 115886252
Conc: 58.85 ng/ml




#15 Picloram

R.T.: 11.007 min
Delta R.T.: -0.026 min
Response: 62407647
Conc: 20.14 ng/ml




#15 Picloram

R.T.: 11.012 min
Delta R.T.: 0.000 min
Response: 9661531
Conc: 2.84 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 11.484 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 10.961 min
Delta R.T.: -0.002 min
Response: 5195911
Conc: 1.55 ng/ml