

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041822\
 Data File : PS019029.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2022 15:51
 Operator : AJ\MA
 Sample : HSTDICV750
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ICVPS041822

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 03:38:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041822.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 18 17:28:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.221	6.799	2632.6E6	531.5E6	703.891	732.219
Target Compounds							
1) T	Dalapon	3.026	2.529	3046.8E6	622.0E6	678.833	669.970
2) T	3,5-DICHL...	6.431	5.975	3349.3E6	768.0E6	673.562	685.491
3) T	4-Nitroph...	7.056	6.414	752.5E6	268.7E6	574.604	676.013
5) T	DICAMBA	7.412	6.965	10165.4E6	2268.6E6	736.558	701.139
6) T	MCP	7.583	7.063	827.1E6	179.0E6	75.919	69.629
7) T	MCPA	7.733	7.277	1114.8E6	249.3E6	73.977	68.641
8) T	DICHLORPROP	8.106	7.587	2541.3E6	577.9E6	713.492	691.561
9) T	2,4-D	8.337	7.879	2945.1E6	664.3E6	717.625	699.886
10) T	Pentachlo...	8.645	8.307	33822.8E6	7562.4E6	733.499	699.609
11) T	2,4,5-TP ...	9.197	8.668	13940.9E6	3320.0E6	700.054	705.093
12) T	2,4,5-T	9.482	9.047	14006.7E6	3072.8E6	704.873	711.550
13) T	2,4-DB	10.036	9.553	2510.7E6	393.8E6	712.919	710.458
14) T	DINOSEB	11.195	9.880	10469.5E6	2340.9E6	684.250	698.183
15) T	Picloram	11.017	10.861	19200.4E6	4340.2E6	716.522	733.023
16) T	DCPA	11.462	10.810	15893.0E6	4069.2E6	692.638	710.945

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041822\
Data File : PS019029.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2022 15:51
Operator : AJ\MA
Sample : HSTDICV750
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
ICVPS041822

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 03:38:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041822.M
Quant Title : 8080.M
QLast Update : Mon Apr 18 17:28:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

