

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041822\
 Data File : PS019055.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Apr 2022 06:05
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 06:42:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041822.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 18 17:28:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.220	6.798	2767.8E6	584.4E6	740.051	805.132
Target Compounds							
1) T	Dalapon	3.026	2.530	3054.7E6	647.5E6	680.602	697.356
2) T	3,5-DICHL...	6.430	5.975	3616.0E6	822.6E6	727.192	734.251
3) T	4-Nitroph...	7.049	6.413	1315.4E6	301.6E6	1004.433	758.740
5) T	DICAMBA	7.412	6.965	10192.1E6	2401.1E6	738.492	742.107
6) T	MCP	7.583	7.064	815.8E6	189.5E6	74.882	73.717
7) T	MCPA	7.733	7.277	1107.6E6	266.9E6	73.496	73.491
8) T	DICHLORPROP	8.105	7.587	2529.8E6	619.4E6	710.277	741.276
9) T	2,4-D	8.334	7.878	3027.9E6	736.1E6	737.796	775.575
10) T	Pentachlo...	8.645	8.307	32747.4E6	7935.5E6	710.179	734.118
11) T	2,4,5-TP ...	9.196	8.669	13692.1E6	3530.2E6	687.558	749.749
12) T	2,4,5-T	9.480	9.046	13679.2E6	3357.4E6	688.394	777.461
13) T	2,4-DB	10.033	9.553	2786.5E6	446.0E6	791.228	804.656
14) T	DINOSEB	11.195	9.880	10294.6E6	2446.6E6	672.818	729.712
15) T	Picloram	11.014	10.860	20103.2E6	4551.6E6	750.215	768.728
16) T	DCPA	11.462	10.812	16828.1E6	4292.5E6	733.394	749.959

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041822\
Data File : PS019055.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Apr 2022 06:05
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 06:42:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041822.M
Quant Title : 8080.M
QLast Update : Mon Apr 18 17:28:16 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

