

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041919\
 Data File : PS004169.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Apr 2019 17:42
 Operator : EI\SJ
 Sample : MDL-HERB-WATER-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MDL-HERB-WATER-02

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 25 02:19:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 11 14:18:42 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.222	10.728	244.2E6	213.6E6	503.344m	465.208
Target Compounds							
1) T	Dalapon	2.845	2.332	63967121	62004866	93.370m	98.663m
2) T	3,5-DICHL...	9.673	8.904	67651521	59554375	94.535m	104.184
3) T	4-Nitroph...	11.009	9.937	7117168	17390187	28.176m	72.517 #
5) T	DICAMBA	11.521	11.005	190.8E6	171.3E6	97.287m	104.072m
6) T	MCPD	11.853	11.227	9024429	7968455	6.913	7.640m
7) T	MCPA	12.103	11.602	15446121	13198310	7.487m	7.254m
8) T	DICHLORPROP	12.720	12.135	48428125	47584616	100.038m	112.091m
9) T	2,4-D	13.132	12.623	24260282	51808347	53.281m	101.245 #
10) T	Pentachlo...	13.428	13.130	752.9E6	668.9E6	103.805m	107.702
11) T	2,4,5-TP ...	14.315	13.765	267.6E6	249.6E6	98.828m	103.653
12) T	2,4,5-T	14.743	14.338	276.2E6	234.4E6	93.805m	101.609m
13) T	2,4-DB	15.495	14.990	16985115	23619064	54.541m	76.664m#
14) T	DINOSEB	16.911	15.395	230.8E6	188.6E6	92.128m	100.912m
15) T	Picloram	16.735	16.644	211.5E6	274.9E6	58.974m	80.294 #
16) T	DCPA	17.259	16.569	357.7E6	335.6E6	99.361m	105.602

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041919\
Data File : PS004169.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Apr 2019 17:42
Operator : EI\SJ
Sample : MDL-HERB-WATER-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
MDL-HERB-WATER-02

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 25 02:19:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041119.M
Quant Title : 8080.M
QLast Update : Thu Apr 11 14:18:42 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

