

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041924\
 Data File : PS026475.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Apr 2024 10:49
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 12:15:05 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041924.M
 Quant Title : 8080.M
 QLast Update : Fri Apr 19 12:09:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.316	7.803	564.7E6	393.2E6	218.547	213.679
Target Compounds							
1) T	Dalapon	2.674	2.717	752.8E6	528.6E6	194.244	183.754
2) T	3,5-DICHL...	6.476	6.744	699.3E6	492.5E6	199.302	196.514
3) T	4-Nitroph...	7.118	7.325	324.1E6	232.8E6	194.441	190.130
5) T	DICAMBA	7.506	8.005	2040.4E6	1547.0E6	196.712	191.454
6) T	MCP	7.686	8.102	128.0E6	99659669	17.820	18.122
7) T	MCPA	7.835	8.348	203.7E6	153.6E6	18.844	18.796
8) T	DICHLORPROP	8.218	8.723	636.3E6	439.4E6	206.626	201.758
9) T	2,4-D	8.449	9.054	690.7E6	482.2E6	203.330	201.794
10) T	Pentachlo...	8.751	9.586	8952.8E6	6420.8E6	201.928	202.496
11) T	2,4,5-TP ...	9.333	9.960	3574.9E6	2546.0E6	202.889	201.842
12) T	2,4,5-T	9.627	10.381	3714.2E6	2575.3E6	202.262	202.363
13) T	2,4-DB	10.206	10.947	520.4E6	334.2E6	202.198	202.778
14) T	DINOSEB	11.426	11.328	2335.9E6	1708.8E6	200.615	200.886
15) T	Picloram	11.230	12.426	4754.0E6	3639.5E6	198.631	193.770
16) T	DCPA	11.716	12.372	4198.5E6	3219.6E6	205.142	205.702

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041924\
Data File : PS026475.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Apr 2024 10:49
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 12:15:05 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041924.M
Quant Title : 8080.M
QLast Update : Fri Apr 19 12:09:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

