

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042020\
 Data File : PS009875.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2020 20:32
 Operator : DD\AJ
 Sample : L2249-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 00:52:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.236	6.922	274.7E6	247.9E6	454.989	500.647
Target Compounds							
1) T	Dalapon	0.000	2.583	0	3744054	N.D.	5.230
2) T	3,5-DICHL...	0.000	6.102	0	65864685	N.D.	88.658
3) T	4-Nitroph...	7.010	6.513	8063331	1707083	28.984	5.621 #
5) T	DICAMBA	7.462	7.098	12400663	3690761	5.505	1.856 #
6) T	MCPD	7.651	7.204	5290594	680086	2.963	<MDL #
7) T	MCPA	7.785	7.388	9913860	9837932	3.846	4.124
8) T	DICHLORPROP	8.071	7.742	18132616	10418560	32.672	20.700 #
9) T	2,4-D	8.417	0.000	19324862	0	33.656	N.D. #
10) T	Pentachlo...	0.000	8.516	0	2071447	N.D.	<MDL
11) T	2,4,5-TP ...	0.000	8.848	0	5819261	N.D.	2.225
12) T	2,4,5-T	9.442	9.239	37380083	5899202	13.482	2.261 #
13) T	2,4-DB	0.000	9.710	0	1202249	N.D.	3.391
14) T	DINOSEB	0.000	10.038	0	9794812	N.D.	4.974
15) T	Picloram	0.000	11.009	0	7842110	N.D.	2.306
16) T	DCPA	0.000	10.953	0	6679886	N.D.	1.992



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042020\
Data File : PS009875.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2020 20:32
Operator : DD\AJ
Sample : L2249-05
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 00:52:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.040 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.583 min
Delta R.T.: -0.039 min
Response: 3744054
Conc: 5.23 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 6.448 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.102 min
Delta R.T.: 0.027 min
Response: 65864685
Conc: 88.66 ng/ml



#3 4-Nitrophenol


R.T.: 7.010 min
Delta R.T.: -0.058 min
Response: 8063331
Conc: 28.98 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.513 min
Delta R.T.: -0.018 min
Response: 1707083
Conc: 5.62 ng/ml




#4 2,4-DCAA

R.T.: 7.236 min
Delta R.T.: -0.003 min
Response: 274729511
Conc: 454.99 ng/ml



#4 2,4-DCAA


R.T.: 6.922 min
Delta R.T.: -0.006 min
Response: 247945540
Conc: 500.65 ng/ml



#5 DICAMBA


R.T.: 7.462 min
Delta R.T.: 0.030 min
Response: 12400663
Conc: 5.51 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 3690761
Conc: 1.86 ng/ml




#6 MCP

R.T.: 7.651 min
Delta R.T.: 0.049 min
Response: 5290594
Conc: 2.96 ug/ml



#6 MCP


R.T.: 7.204 min
Delta R.T.: 0.008 min
Response: 680086
Conc: N.D.



#7 MCPA


R.T.: 7.785 min
Delta R.T.: 0.034 min
Response: 9913860
Conc: 3.85 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.388 min
Delta R.T.: -0.023 min
Response: 9837932
Conc: 4.12 ug/ml




#8 DICHLORPROP

R.T.: 8.071 min
Delta R.T.: -0.055 min
Response: 18132616
Conc: 32.67 ng/ml



#8 DICHLORPROP


R.T.: 7.742 min
Delta R.T.: 0.017 min
Response: 10418560
Conc: 20.70 ng/ml



#9 2,4-D


R.T.: 8.417 min
Delta R.T.: 0.063 min
Response: 19324862
Conc: 33.66 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.019 min
Response: 0
Conc: N.D.




#11 2,4,5-TP (SILVEX)

R.T.: 0.000 min
Exp R.T. : 9.216 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 8.848 min
Delta R.T.: 0.036 min
Response: 5819261
Conc: 2.23 ng/ml



#12 2,4,5-T


R.T.: 9.442 min
Delta R.T.: -0.059 min
Response: 37380083
Conc: 13.48 ng/ml

Instrument :
ECD_S
ClientSampleId :




#12 2,4,5-T

R.T.: 9.239 min
Delta R.T.: 0.047 min
Response: 5899202
Conc: 2.26 ng/ml




#13 2,4-DB

R.T.: 0.000 min
Exp R.T. : 10.053 min
Response: 0
Conc: N.D.



#13 2,4-DB


R.T.: 9.710 min
Delta R.T.: 0.011 min
Response: 1202249
Conc: 3.39 ng/ml



#14 DINOSEB


R.T.: 0.000 min
Exp R.T. : 11.219 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 10.038 min
Delta R.T.: 0.004 min
Response: 9794812
Conc: 4.97 ng/ml




#15 Picloram

R.T.: 0.000 min
Exp R.T. : 11.033 min
Response: 0
Conc: N.D.



#15 Picloram


R.T.: 11.009 min
Delta R.T.: -0.004 min
Response: 7842110
Conc: 2.31 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 11.484 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 10.953 min
Delta R.T.: -0.011 min
Response: 6679886
Conc: 1.99 ng/ml