

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042020\
 Data File : PS009880.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2020 22:27
 Operator : DD\AJ
 Sample : PB128303BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 00:53:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	316.5E6	307.4E6	524.102	620.631
Target Compounds							
2) T	3,5-DICHL...	0.000	6.101	0	106.8E6	N.D.	143.728
3) T	4-Nitroph...	7.010	6.513	8574878	4315688	30.823	14.211 #
5) T	DICAMBA	0.000	7.099	0	3694185	N.D.	1.857
6) T	MCPD	0.000	7.206	0	1465618	N.D.	<MDL
7) T	MCPA	7.786	7.388	34472927	32827068	13.374	13.762
8) T	DICHLORPROP	0.000	7.740	0	29430926	N.D.	58.474
10) T	Pentachlo...	0.000	8.395	0	930095	N.D.	<MDL
11) T	2,4,5-TP ...	9.279	8.848	20425065	41500139	6.902	15.869 #
12) T	2,4,5-T	9.521	9.190	10575403	3276351	3.814	1.256 #
13) T	2,4-DB	10.092	9.709	15338511	1966489	40.492	5.546 #
14) T	DINOSEB	0.000	10.034	0	47902069	N.D.	24.325
15) T	Picloram	11.030	11.006	281.0E6	292.1E6	90.687	85.872
16) T	DCPA	0.000	10.955	0	7078781	N.D.	2.111



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042020\
Data File : PS009880.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2020 22:27
Operator : DD\AJ
Sample : PB128303BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 00:53:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#2 3,5-DICHLOROBENZOIC ACID


R.T.: 0.000 min
Exp R.T.: 6.448 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.101 min
Delta R.T.: 0.027 min
Response: 106776487
Conc: 143.73 ng/ml




#3 4-Nitrophenol

R.T.: 7.010 min
Delta R.T.: -0.057 min
Response: 8574878
Conc: 30.82 ng/ml



#3 4-Nitrophenol


R.T.: 6.513 min
Delta R.T.: -0.017 min
Response: 4315688
Conc: 14.21 ng/ml



#4 2,4-DCAA


R.T.: 7.235 min
Delta R.T.: -0.004 min
Response: 316461001
Conc: 524.10 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.922 min
Delta R.T.: -0.006 min
Response: 307367357
Conc: 620.63 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.432 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 7.099 min
Delta R.T.: 0.000 min
Response: 3694185
Conc: 1.86 ng/ml



#7 MCPA


R.T.: 7.786 min
Delta R.T.: 0.035 min
Response: 34472927
Conc: 13.37 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.388 min
Delta R.T.: -0.024 min
Response: 32827068
Conc: 13.76 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 8.125 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.740 min
Delta R.T.: 0.015 min
Response: 29430926
Conc: 58.47 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.279 min
Delta R.T.: 0.063 min
Response: 20425065
Conc: 6.90 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.848 min
Delta R.T.: 0.036 min
Response: 41500139
Conc: 15.87 ng/ml




#12 2,4,5-T

R.T.: 9.521 min
Delta R.T.: 0.020 min
Response: 10575403
Conc: 3.81 ng/ml



#12 2,4,5-T


R.T.: 9.190 min
Delta R.T.: -0.001 min
Response: 3276351
Conc: 1.26 ng/ml



#13 2,4-DB


R.T.: 10.092 min
Delta R.T.: 0.038 min
Response: 15338511
Conc: 40.49 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 9.709 min
Delta R.T.: 0.010 min
Response: 1966489
Conc: 5.55 ng/ml




#14 DINOSEB

R.T.: 0.000 min
Exp R.T. : 11.219 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 10.034 min
Delta R.T.: 0.000 min
Response: 47902069
Conc: 24.32 ng/ml



#15 Picloram


R.T.: 11.030 min
Delta R.T.: -0.003 min
Response: 280953900
Conc: 90.69 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 11.006 min
Delta R.T.: -0.007 min
Response: 292068695
Conc: 85.87 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T. : 11.484 min
Response: 0
Conc: N.D.



#16 DCPA

R.T.: 10.955 min
Delta R.T.: -0.009 min
Response: 7078781
Conc: 2.11 ng/ml