

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042021\
 Data File : PS014709.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2021 16:42
 Operator : DD\AJ
 Sample : M2075-28MS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-8-DMS

Manual Integrations
 APPROVED

Ankita
 4/22/2021 11:53:53 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 18:04:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.156	6.882	577.3E6	207.6E6	379.493	446.105
Target Compounds							
1) T	Dalapon	3.007	2.605	506.9E6	226.4E6	281.093	347.853m
2) T	3,5-DICHL...	6.382	6.045	700.4E6	232.8E6	338.053	317.499
5) T	DICAMBA	7.345	7.050	1956.9E6	695.5E6	336.331	343.897
6) T	MCP	7.509	7.142	136.3E6	52255967	31.293	31.030
7) T	MCPA	7.657	7.354	203.1E6	107.5E6	32.638	44.280m#
8) T	DICHLORPROP	8.031	7.712	477.0E6	2173.3E6	340.842	4305.127 #
9) T	2,4-D	8.260	7.970	646.7E6	247.6E6	406.714	445.161
10) T	Pentachlo...	8.567	8.395	2756.0E6	922.1E6	131.132	137.172
11) T	2,4,5-TP ...	9.116	8.755	2710.2E6	897.3E6	331.064	328.469
12) T	2,4,5-T	9.398	9.143	2557.7E6	1499.0E6	310.922	573.108 #
13) T	2,4-DB	9.984f	9.660	41069.6E6	14465.6E6	26988.314	41702.961m#
14) T	DINOSEB	11.060	9.981	1044.7E6	939.4E6	162.135m	487.848 #
15) T	Picloram	10.923	10.956	2636.1E6	1134.1E6	235.611	302.433 #
16) T	DCPA	11.360	10.897	3719.3E6	1698.0E6	301.844	418.156 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042021\
Data File : PS014709.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2021 16:42
Operator : DD\AJ
Sample : M2075-28MS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
TP-8-DMS

Manual Integrations
APPROVED

Ankita
4/22/2021 11:53:53 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 18:04:20 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

