

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042022\
 Data File : PS019100.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2022 23:45
 Operator : AJ\MA
 Sample : N2468-04MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-15-1-6MS

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 04/22/2022
 Supervised By : mohammad ahmed 04/22/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 01:06:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041822.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 18 17:28:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.216	6.796	1985.9E6	436.9E6	530.981	601.952
Target Compounds							
1) T	Dalapon	3.025	2.535	1945.1E6	516.5E6	433.371	556.325 #
2) T	3,5-DICHL...	6.428	5.973	2565.1E6	546.5E6	515.850	487.776m
5) T	DICAMBA	7.408	6.963	6916.2E6	1628.9E6	501.131	503.429
6) T	MCP	7.578	7.060	590.4E6	141.3E6	54.192	54.948
7) T	MCPA	7.727	7.272	857.7E6	175.2E6	56.914	48.245
8) T	DICHLORPROP	8.101	7.585	1930.6E6	455.6E6	542.024	545.193
9) T	2,4-D	8.329	7.876	2322.0E6	571.3E6	565.788	601.974
10) T	Pentachlo...	8.641	8.305	24066.9E6	5637.2E6	521.928	521.503
11) T	2,4,5-TP ...	9.192	8.666	10828.9E6	2895.5E6	543.782	614.952
12) T	2,4,5-T	9.480	9.047	9890.7E6	2397.1E6	497.738	555.090
13) T	2,4-DB	10.030	9.557	1504.7E6	660.7E6	427.255	1191.945 #
14) T	DINOSEB	11.190	9.878	6926.3E6	1675.5E6	452.681	499.716
15) T	Picloram	11.010	10.858	13142.2E6	3109.7E6	490.441	525.199
16) T	DCPA	11.458	10.809	11284.9E6	3221.2E6	491.812	562.787

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042022\
 Data File : PS019100.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2022 23:45
 Operator : AJ\MA
 Sample : N2468-04MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-15-1-6MS

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 04/22/2022
 Supervised By : mohammad ahmed 04/22/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 01:06:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041822.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 18 17:28:16 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

