

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042023\
 Data File : PS022785.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 16:50
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/21/2023
 Supervised By :Ankita Jodhani 04/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 21:39:41 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041223.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 12 14:05:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.485	6.864	1924.3E6	804.7E6	772.903	789.716
Target Compounds							
1) T	Dalapon	2.545	2.572	2057.1E6	1115.7E6	696.533	704.771
2) T	3,5-DICHL...	5.837	6.028	2512.7E6	1155.5E6	743.613	740.995
3) T	4-Nitroph...	6.316	6.468	1271.2E6	463.1E6	685.061m	728.413
5) T	DICAMBA	6.634	7.032	8287.9E6	3529.7E6	754.694	766.626
6) T	MCPD	6.797	7.131	596.2E6	266.4E6	70.856	76.483
7) T	MCPA	6.927	7.342	861.3E6	358.5E6	72.560	74.154
8) T	DICHLORPROP	7.245	7.655	2161.8E6	849.8E6	710.715	763.149
9) T	2,4-D	7.446	7.944	2607.3E6	990.2E6	704.107	773.118
10) T	Pentachlo...	7.681	8.377	29737.5E6	12279.5E6	743.490	765.508
11) T	2,4,5-TP ...	8.217	8.737	13096.5E6	4826.7E6	714.540	774.041
12) T	2,4,5-T	8.477	9.112	13043.0E6	4577.9E6	697.433	773.788
13) T	2,4-DB	8.992	9.619	2855.8E6	573.5E6	668.318	765.461
14) T	DINOSEB	10.068	9.950	9308.7E6	3190.3E6	656.663	722.550
15) T	Picloram	9.887	10.919	19968.7E6	6027.4E6	670.406	741.038
16) T	DCPA	10.335	10.880	15840.2E6	5206.9E6	708.563	749.809

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042023\
Data File : PS022785.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 16:50
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

HSTDCCC750

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 04/21/2023

Supervised By :Ankita Jodhani 04/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 21:39:41 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041223.M
Quant Title : 8080.M
QLast Update : Wed Apr 12 14:05:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

