

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
 Data File : PS009905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2020 11:23
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 15:31:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.237	6.922	340.5E6	335.1E6	563.898	676.713
Target Compounds							
5) T	DICAMBA	0.000	7.157	0	3880957	N.D.	1.951
6) T	MCPD	0.000	7.157	0	3880957	N.D.	2.358
8) T	DICHLORPROP	0.000	7.744	0	3480069	N.D.	6.914
9) T	2,4-D	8.422	0.000	13111844	0	22.836	N.D. #
11) T	2,4,5-TP ...	0.000	8.805	0	1145985	N.D.	<MDL
12) T	2,4,5-T	0.000	9.186	0	1145821	N.D.	<MDL
15) T	Picloram	11.033	11.007	10232533	15734644	3.303	4.626 #
16) T	DCPA	0.000	10.955	0	1751878	N.D.	<MDL



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
Data File : PS009905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2020 11:23
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 15:31:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 7.237 min
Delta R.T.: -0.002 min
Response: 340490071
Conc: 563.90 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 6.922 min
Delta R.T.: -0.006 min
Response: 335141915
Conc: 676.71 ng/ml




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.432 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 7.157 min
Delta R.T.: 0.059 min
Response: 3880957
Conc: 1.95 ng/ml



#6 MCP


R.T.: 0.000 min
Exp R.T.: 7.602 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#6 MCP

R.T.: 7.157 min
Delta R.T.: -0.039 min
Response: 3880957
Conc: 2.36 ug/ml




#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T.: 8.125 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 7.744 min
Delta R.T.: 0.019 min
Response: 3480069
Conc: 6.91 ng/ml



#9 2,4-D


R.T.: 8.422 min
Delta R.T.: 0.068 min
Response: 13111844
Conc: 22.84 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 8.019 min
Response: 0
Conc: N.D.



#15 Picloram

R.T.: 11.033 min
Delta R.T.: 0.000 min
Response: 10232533
Conc: 3.30 ng/ml



#15 Picloram

R.T.: 11.007 min
Delta R.T.: -0.006 min
Response: 15734644
Conc: 4.63 ng/ml