

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
 Data File : PS009937.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2020 01:02
 Operator : DD\AJ
 Sample : L2327-06
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 03:03:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	237.7E6	243.3E6	393.617	491.212
Target Compounds							
1) T	Dalapon	0.000	2.582	0	12414148	N.D.	17.341
2) T	3,5-DICHL...	0.000	6.101	0	93415595	N.D.	125.743
3) T	4-Nitroph...	0.000	6.512	0	3881926	N.D.	12.783
5) T	DICAMBA	7.417	7.056	7859870	23478447	3.489	11.804 #
6) T	MCPD	0.000	7.224	0	1607668	N.D.	<MDL
7) T	MCPA	0.000	7.471	0	39897172	N.D.	16.726
8) T	DICHLORPROP	0.000	7.740	0	27825083	N.D.	55.283
9) T	2,4-D	8.420	0.000	6421496	0	11.184	N.D. #
10) T	Pentachlo...	0.000	8.438	0	3848547	N.D.	<MDL
11) T	2,4,5-TP ...	0.000	8.846	0	7902249	N.D.	3.022
12) T	2,4,5-T	9.441	9.255	34835449	8391749	12.564	3.216 #
13) T	2,4-DB	0.000	9.706	0	1402925	N.D.	3.957
14) T	DINOSEB	0.000	10.071	0	30230894	N.D.	15.351
15) T	Picloram	11.006	11.006	20917078	4923836	6.752	1.448 #
16) T	DCPA	0.000	10.958	0	1379172	N.D.	<MDL



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
Data File : PS009937.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2020 01:02
Operator : DD\AJ
Sample : L2327-06
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 03:03:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.040 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.582 min
Delta R.T.: -0.040 min
Response: 12414148
Conc: 17.34 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T.: 6.448 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.101 min
Delta R.T.: 0.027 min
Response: 93415595
Conc: 125.74 ng/ml



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T.: 7.068 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.512 min
Delta R.T.: -0.019 min
Response: 3881926
Conc: 12.78 ng/ml




#4 2,4-DCAA

R.T.: 7.235 min
Delta R.T.: -0.004 min
Response: 237672241
Conc: 393.62 ng/ml



#4 2,4-DCAA


R.T.: 6.922 min
Delta R.T.: -0.006 min
Response: 243272986
Conc: 491.21 ng/ml



#5 DICAMBA


R.T.: 7.417 min
Delta R.T.: -0.015 min
Response: 7859870
Conc: 3.49 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 7.056 min
Delta R.T.: -0.042 min
Response: 23478447
Conc: 11.80 ng/ml




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 7.751 min
Response: 0
Conc: N.D.



#7 MCPA


R.T.: 7.471 min
Delta R.T.: 0.060 min
Response: 39897172
Conc: 16.73 ug/ml



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 8.125 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.740 min
Delta R.T.: 0.015 min
Response: 27825083
Conc: 55.28 ng/ml




#9 2,4-D

R.T.: 8.420 min
Delta R.T.: 0.066 min
Response: 6421496
Conc: 11.18 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.019 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 0.000 min
Exp R.T.: 9.216 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.846 min
Delta R.T.: 0.034 min
Response: 7902249
Conc: 3.02 ng/ml




#12 2,4,5-T

R.T.: 9.441 min
Delta R.T.: -0.059 min
Response: 34835449
Conc: 12.56 ng/ml



#12 2,4,5-T


R.T.: 9.255 min
Delta R.T.: 0.064 min
Response: 8391749
Conc: 3.22 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T.: 10.053 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 9.706 min
Delta R.T.: 0.008 min
Response: 1402925
Conc: 3.96 ng/ml




#14 DINOSEB

R.T.: 0.000 min
Exp R.T.: 11.219 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 10.071 min
Delta R.T.: 0.037 min
Response: 30230894
Conc: 15.35 ng/ml



#15 Picloram

R.T.: 11.006 min
Delta R.T.: -0.026 min
Response: 20917078
Conc: 6.75 ng/ml

Instrument :
ECD_S
ClientSampleId :



#15 Picloram

R.T.: 11.006 min
Delta R.T.: -0.006 min
Response: 4923836
Conc: 1.45 ng/ml