

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
 Data File : PS009924.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2020 20:05
 Operator : DD\AJ
 Sample : L2244-12MSD
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-23-22-COMPMSD

Manual Integrations
 APPROVED

Sohil
 4/22/2020 12:05:57 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 02:59:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	222.0E6	228.2E6	367.675m	460.719 #
Target Compounds							
1) T	Dalapon	3.043	2.628	324.4E6	352.9E6	428.216	492.970
2) T	3,5-DICHL...	6.446	6.071	397.2E6	369.3E6	483.054	497.082
5) T	DICAMBA	7.428	7.092	1224.0E6	1145.1E6	543.359	575.729
6) T	MCP	7.597	7.188	93904859	106.8E6	52.591	64.858
7) T	MCPA	7.746	7.403	148.2E6	138.0E6	57.498m	57.839
8) T	DICHLORPROP	8.122	7.718	318.1E6	325.7E6	573.158	647.123
9) T	2,4-D	8.351	8.011	384.6E6	438.2E6	669.742	731.216
10) T	Pentachlo...	8.664	8.443	3321.5E6	3062.6E6	446.987	490.796
11) T	2,4,5-TP ...	9.214	8.802	1848.2E6	2016.1E6	624.498	770.896m
12) T	2,4,5-T	9.498	9.183	1706.4E6	1707.3E6	615.421	654.262
13) T	2,4-DB	10.051	9.690	205.2E6	225.0E6	541.694	634.593
14) T	DINOSEB	11.216	10.025	804.8E6	792.9E6	359.657	402.660
15) T	Picloram	11.030	11.004	865.1E6	912.3E6	279.239	268.231
16) T	DCPA	11.480	10.954	2136.0E6	2116.5E6	573.618	631.232

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
Data File : PS009924.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2020 20:05
Operator : DD\AJ
Sample : L2244-12MSD
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-23-22-COMPMSD

Manual Integrations
APPROVED

Sohil
4/22/2020 12:05:57 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 02:59:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

