

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
 Data File : PS009942.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2020 02:57
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 04:48:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	501.4E6	429.1E6	830.314	866.529
Target Compounds							
1) T	Dalapon	3.045	2.628	454.6E6	445.7E6	600.102	622.535
2) T	3,5-DICHL...	6.447	6.071	621.3E6	590.3E6	755.500	794.555
3) T	4-Nitroph...	7.065	6.526	211.6E6	238.1E6	760.549	783.888
5) T	DICAMBA	7.428	7.091	1774.7E6	1616.9E6	787.839	812.916
6) T	MCPD	7.599	7.190	143.1E6	137.3E6	80.120	83.418
7) T	MCPA	7.749	7.406	199.7E6	194.7E6	77.469	81.608
8) T	DICHLORPROP	8.122	7.718	427.8E6	419.5E6	770.814	833.546
9) T	2,4-D	8.351	8.012	447.8E6	488.5E6	779.937	815.169
10) T	Pentachlo...	8.664	8.445	5831.7E6	5080.0E6	784.794	814.099
11) T	2,4,5-TP ...	9.213	8.804	2373.5E6	2190.2E6	802.001	837.496
12) T	2,4,5-T	9.497	9.183	2231.7E6	2164.6E6	804.902	829.478
13) T	2,4-DB	10.051	9.690	303.2E6	294.5E6	800.410	830.468
14) T	DINOSEB	11.215	10.025	1677.9E6	1597.5E6	749.863	811.204
15) T	Picloram	11.030	11.004	2269.9E6	2268.6E6	732.693	667.001
16) T	DCPA	11.480	10.954	2937.3E6	2771.7E6	788.830	826.629

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042120\
Data File : PS009942.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2020 02:57
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 04:48:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

