

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042321\
 Data File : PS014775.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2021 08:45
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 4/27/2021 10:51:26 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 10:28:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.156	6.883	1212.5E6	341.1E6	797.085	732.872
Target Compounds							
1) T	Dalapon	3.008	2.606	1189.7E6	411.6E6	659.781	632.393
2) T	3,5-DICHL...	6.383	6.045	1549.9E6	542.1E6	748.019	739.298
3) T	4-Nitroph...	6.989	6.497	623.1E6	221.5E6	766.807	784.673m
5) T	DICAMBA	7.345	7.049	4414.7E6	1507.3E6	758.762	745.258
6) T	MCPD	7.513	7.146	336.8E6	130.9E6	77.347	77.758
7) T	MCPA	7.661	7.361	470.9E6	178.9E6	75.670	73.659
8) T	DICHLORPROP	8.031	7.674	1066.8E6	379.3E6	762.342	751.428
9) T	2,4-D	8.259	7.969	1200.7E6	447.2E6	755.191	804.221
10) T	Pentachlo...	8.566	8.394	15824.4E6	5037.1E6	752.944	749.284
11) T	2,4,5-TP ...	9.110	8.755	6220.3E6	2060.7E6	759.831	754.371
12) T	2,4,5-T	9.394	9.134	6257.3E6	2038.2E6	760.661	779.278
13) T	2,4-DB	9.941	9.639	1211.6E6	267.7E6	796.215	771.712
14) T	DINOSEB	11.094	9.970	4704.6E6	1443.4E6	730.116	749.539
15) T	Picloram	10.918	10.952	8405.8E6	2814.9E6	751.303	750.649
16) T	DCPA	11.359	10.896	9183.6E6	3082.7E6	745.302	759.170

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042321\
Data File : PS014775.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2021 08:45
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
4/27/2021 10:51:26 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 10:28:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

