

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042324\
 Data File : PS026527.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Apr 2024 15:11
 Operator : AR\AJ
 Sample : P2208-01MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LNG-11MSD

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 04/24/2024
 Supervised By :Ankita Jodhani 04/24/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 03:19:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041924.M
 Quant Title : 8080.M
 QLast Update : Fri Apr 19 12:45:38 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.306	7.800	2904.0E6	1483.0E6	1180.295m	836.451 #
Target Compounds							
1) T	Dalapon	2.671	2.715	1420.0E6	1502.1E6	374.271	530.438 #
2) T	3,5-DICHL...	6.466	6.743	1544.4E6	1182.3E6	457.900m	487.280m
5) T	DICAMBA	7.496	8.001	4310.2E6	3253.2E6	427.919m	410.681m
6) T	MCP	7.677	8.101	290.6E6	252.0E6	39.574m	45.314
7) T	MCPA	7.827	8.346	497.5E6	325.6E6	46.297	40.132m
8) T	DICHLORPROP	8.206	8.719	1295.4E6	858.9E6	441.218m	409.196
9) T	2,4-D	8.438	9.051	1753.4E6	1182.4E6	539.023	514.544
10) T	Pentachlo...	8.738	9.583	20152.1E6	13732.0E6	490.637	457.692
11) T	2,4,5-TP ...	9.321	9.958	8694.1E6	5906.9E6	517.088m	488.517m
12) T	2,4,5-T	9.621	10.380	8343.9E6	5278.7E6	475.644	432.323
13) T	2,4-DB	10.192	10.945	1037.6E6	747.0E6	415.472m	467.547m
14) T	DINOSEB	11.409	11.324	5020.9E6	4351.9E6	447.906m	529.863m
15) T	Picloram	11.214	12.423	10047.4E6	8023.6E6	432.918m	437.690
16) T	DCPA	11.700	12.368	8672.4E6	7287.5E6	445.396	489.939

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042324\
Data File : PS026527.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Apr 2024 15:11
Operator : AR\AJ
Sample : P2208-01MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :

ECD_S

ClientSampleId :

LNG-11MSD

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 04/24/2024

Supervised By :Ankita Jodhani 04/24/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 03:19:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041924.M
Quant Title : 8080.M
QLast Update : Fri Apr 19 12:45:38 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

