

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042523\
 Data File : PS022877.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2023 21:44
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 01:06:52 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041223.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 12 14:05:56 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.484	6.858	1881.8E6	759.8E6	755.852	745.658
Target Compounds							
1) T	Dalapon	2.543	2.569	1919.9E6	1045.2E6	650.084	660.250
2) T	3,5-DICHL...	5.836	6.023	2380.0E6	1073.9E6	704.336	688.661
3) T	4-Nitroph...	6.316	6.463	1262.2E6	431.1E6	680.201	678.086
5) T	DICAMBA	6.633	7.026	8029.8E6	3306.1E6	731.190	718.066
6) T	MCP	6.796	7.124	632.6E6	250.6E6	75.178	71.967
7) T	MCPA	6.926	7.335	863.3E6	335.0E6	72.721	69.275
8) T	DICHLORPROP	7.245	7.648	2177.6E6	795.6E6	715.920	714.466
9) T	2,4-D	7.446	7.936	2637.7E6	923.3E6	712.323	720.903
10) T	Pentachlo...	7.681	8.369	30538.7E6	11576.3E6	763.520	721.674
11) T	2,4,5-TP ...	8.218	8.729	13330.2E6	4506.9E6	727.289	722.765
12) T	2,4,5-T	8.479	9.104	13375.6E6	4233.9E6	715.220	715.646
13) T	2,4-DB	8.995	9.610	2875.3E6	532.1E6	672.878	710.145
14) T	DINOSEB	10.071	9.941	9477.3E6	2705.7E6	668.559	612.788
15) T	Picloram	9.890	10.909	18943.1E6	5201.4E6	635.972	639.487
16) T	DCPA	10.339	10.872	15740.5E6	4872.8E6	704.100	701.709

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042523\
Data File : PS022877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2023 21:44
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 01:06:52 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS041223.M
Quant Title : 8080.M
QLast Update : Wed Apr 12 14:05:56 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

