

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\
 Method File : PS042919.M
 Title : 8080.M
 Last Update : Tue Apr 30 08:27:54 2019
 Response Via : Initial Calibration

Calibration Files

750 =PS004331.D 1000 =PS004332.D 1500 =PS004333.D
 500 =PS004330.D 200 =PS004329.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	7.812	7.881	7.909	7.846	7.841	7.858	E5	0.48
2) T	3,5-DICHLOROBENZOI...	8.042	7.803	7.789	8.060	7.806	7.900	E5	1.75
3) T	4-Nitrophenol	2.602	2.588	2.684	2.494	2.126	2.499	E5	8.77
4) S	2,4-DCAA	6.134	6.029	6.108	6.264	6.938	6.295	E5	5.87
5) T	DICAMBA	2.370	2.340	2.364	2.362	2.452	2.378	E6	1.82
6) T	MCPP	1.726	1.792	1.906	1.663	1.553	1.728	E6	7.68
7) T	MCPA	2.511	2.540	2.647	2.534	2.651	2.576	E6	2.60
8) T	DICHLORPROP	5.631	5.567	5.739	5.705	5.973	5.723	E5	2.71
9) T	2,4-D	5.709	5.558	5.818	5.811	5.550	5.689	E5	2.30
10) T	Pentachlorophenol	8.565	8.173	7.961	8.939	8.959	8.519	E6	5.26
11) T	2,4,5-TP (SILVEX)	3.302	3.148	3.182	3.314	3.177	3.225	E6	2.39
12) T	2,4,5-T	3.452	3.268	3.328	3.501	3.475	3.405	E6	2.97
13) T	2,4-DB	3.531	3.404	3.780	3.227	2.879	3.364	E5	10.05
14) T	DINOSEB	3.010	2.814	2.701	3.132	3.363	3.004	E6	8.70
15) T	Picloram	4.466	4.325	4.370	4.275	3.729	4.233	E6	6.86
16) T	DCPA	4.428	4.180	3.942	4.635	4.661	4.369	E6	7.03

Signal #2 Calibration Files

750 =PS004331.D 1000 =PS004332.D 1500 =PS004333.D
 500 =PS004330.D 200 =PS004329.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	6.643	6.675	6.696	6.720	6.725	6.692	E5	0.51
2) T	3,5-DICHLOROBENZOI...	6.032	6.049	5.963	6.183	6.508	6.147	E5	3.53
3) T	4-Nitrophenol	2.436	2.401	2.352	2.515	2.664	2.474	E5	4.92
4) S	2,4-DCAA	4.718	4.701	4.719	4.844	5.132	4.823	E5	3.78
5) T	DICAMBA	1.860	1.848	1.808	1.906	1.936	1.872	E6	2.68
6) T	MCPP	1.350	1.372	1.413	1.341	1.248	1.345	E6	4.52
7) T	MCPA	1.950	1.961	1.977	1.985	2.060	1.987	E6	2.17
8) T	DICHLORPROP	4.606	4.590	4.569	4.724	5.039	4.706	E5	4.16
9) T	2,4-D	5.182	5.145	5.139	5.297	5.528	5.258	E5	3.11
10) T	Pentachlorophenol	6.614	6.398	6.010	6.988	7.338	6.670	E6	7.72
11) T	2,4,5-TP (SILVEX)	2.585	2.582	2.517	2.699	2.756	2.627	E6	3.69
12) T	2,4,5-T	2.239	2.207	2.138	2.311	2.351	2.249	E6	3.75
13) T	2,4-DB	2.819	2.873	2.930	2.778	2.307	2.742	E5	9.10
14) T	DINOSEB	1.938	1.923	1.903	1.963	2.046	1.955	E6	2.84
15) T	Picloram	3.664	3.687	3.612	3.685	3.639	3.657	E6	0.88
16) T	DCPA	3.266	3.206	3.088	3.380	3.506	3.289	E6	4.88

(#) = Out of Range