

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042921\
 Data File : PS014870.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2021 01:20
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:15:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.157	6.885	1292.2E6	381.8E6	849.461	820.203
Target Compounds							
1) T	Dalapon	3.007	2.606	1222.9E6	429.6E6	678.195	660.146
2) T	3,5-DICHL...	6.384	6.047	1653.4E6	571.7E6	797.992	779.646
3) T	4-Nitroph...	6.993	6.502	614.6E6	200.7E6	756.371	710.700
5) T	DICAMBA	7.344	7.050	4634.3E6	1609.5E6	796.497	795.775
6) T	MCPD	7.511	7.146	348.8E6	133.6E6	80.113	79.309
7) T	MCPA	7.660	7.361	490.5E6	188.0E6	78.822	77.417
8) T	DICHLORPROP	8.031	7.675	1112.8E6	396.2E6	795.243	784.762
9) T	2,4-D	8.262	7.973	1233.1E6	460.6E6	775.567	828.343
10) T	Pentachlo...	8.565	8.395	16521.6E6	5351.4E6	786.114	796.039
11) T	2,4,5-TP ...	9.110	8.756	6418.9E6	2190.1E6	784.087	801.752
12) T	2,4,5-T	9.396	9.136	6373.6E6	2085.9E6	774.799	797.503
13) T	2,4-DB	9.944	9.642	1179.2E6	273.9E6	774.898	789.597
14) T	DINOSEB	11.092	9.971	5002.1E6	1522.5E6	776.289	790.647
15) T	Picloram	10.922	10.957	7430.8E6	2560.8E6	664.156	682.892
16) T	DCPA	11.358	10.897	9507.2E6	3250.4E6	771.558	800.448

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS042921\
Data File : PS014870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2021 01:20
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:15:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

