

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS043024\
 Data File : PS026588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2024 14:20
 Operator : AR\AJ
 Sample : P2240-02MSD RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 AU-05-042324RE

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 05/01/2024
 Supervised By :mohammad ahmed 05/02/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 03:44:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS042924.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 29 15:59:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.293	7.788	302.5E6	181.3E6	122.258m	107.688
Target Compounds							
1) T	Dalapon	2.665	2.708	606.0E6	705.0E6	162.177m	248.151m#
2) T	3,5-DICHL...	6.456	6.732	280.0E6	186.0E6	83.427	77.123
3) T	4-Nitroph...	7.095	7.312	302.6E6	206.7E6	186.630	180.741
5) T	DICAMBA	7.483	7.990	1216.7E6	906.7E6	119.988	122.245
6) T	MCPD	7.659	8.088	18221359	12031428	2.448	2.262
7) T	MCPA	7.809	8.332	81042385	69791228	7.409	8.880
8) T	DICHLORPROP	8.191	8.707	154.0E6	107.9E6	51.615m	55.496
9) T	2,4-D	8.423	9.038	480.0E6	267.3E6	145.084	128.600
10) T	Pentachlo...	8.722	9.569	1239.2E6	819.7E6	30.256	30.409
11) T	2,4,5-TP ...	9.304	9.938	664.1E6	314.3E6	38.806	28.465m#
12) T	2,4,5-T	9.604	10.367	983.3E6	560.8E6	55.104	51.291m
14) T	DINOSEB	0.000	11.314	0	830.3E6	N.D.	112.864
15) T	Picloram	11.195	12.412	4540.3E6	3921.0E6	190.749m	242.300 #
16) T	DCPA	11.680	12.357	5528.1E6	5191.8E6	281.871	393.308 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS043024\
Data File : PS026588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2024 14:20
Operator : AR\AJ
Sample : P2240-02MSD RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
AU-05-042324RE

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/01/2024
Supervised By : mohammad ahmed 05/02/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 03:44:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS042924.M
Quant Title : 8080.M
QLast Update : Mon Apr 29 15:59:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

