

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS043024\
 Data File : PS026613.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2024 01:58
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 03:48:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS042924.M
 Quant Title : 8080.M
 QLast Update : Mon Apr 29 15:59:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.287	7.792	1900.0E6	1351.3E6	767.954	802.524
Target Compounds							
1) T	Dalapon	2.670	2.718	2521.2E6	2010.3E6	674.761	707.637
2) T	3,5-DICHL...	6.452	6.737	2471.2E6	1853.8E6	736.292	768.796
3) T	4-Nitroph...	7.090	7.316	1151.9E6	847.3E6	710.439	741.055
5) T	DICAMBA	7.477	7.993	7552.7E6	5759.9E6	744.838	776.585
6) T	MCPD	7.658	8.095	490.5E6	410.5E6	65.909	77.193
7) T	MCPA	7.809	8.341	804.8E6	573.3E6	73.577	72.944
8) T	DICHLORPROP	8.184	8.710	2177.3E6	1463.3E6	729.519	752.830
9) T	2,4-D	8.415	9.041	2479.9E6	1578.3E6	749.586	759.429
10) T	Pentachlo...	8.714	9.573	31689.0E6	21013.7E6	773.693	779.599
11) T	2,4,5-TP ...	9.295	9.948	12570.2E6	8490.7E6	734.590	768.921
12) T	2,4,5-T	9.587	10.368	13260.9E6	8465.7E6	743.125	774.302
13) T	2,4-DB	10.163	10.934	1869.6E6	1061.5E6	726.133	737.172
14) T	DINOSEB	11.378	11.314	8253.5E6	5641.7E6	740.775	766.934
15) T	Picloram	11.184	12.412	17009.3E6	12261.6E6	714.598	757.705
16) T	DCPA	11.669	12.359	14243.0E6	10207.9E6	726.240	773.309

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS043024\
Data File : PS026613.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2024 01:58
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 03:48:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS042924.M
Quant Title : 8080.M
QLast Update : Mon Apr 29 15:59:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

