

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050120\
 Data File : PS010107.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2020 19:56
 Operator : DD\AJ
 Sample : L2122-15MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SU-03-043020MS

Manual Integrations
 APPROVED

Sohil
 5/4/2020 8:10:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 03:44:54 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.922	214.4E6	190.3E6	355.119	384.247
Target Compounds							
1) T	Dalapon	3.043	2.628	340.3E6	294.4E6	449.126	411.207m
2) T	3,5-DICHL...	6.446	6.071	402.9E6	347.5E6	489.999	467.700
5) T	DICAMBA	7.427	7.091	1206.2E6	1055.0E6	535.474	530.412
6) T	MCP	7.595	7.187	86896616	96719271	48.667	58.761
7) T	MCPA	7.744	7.402	142.2E6	126.7E6	55.155m	53.114
8) T	DICHLORPROP	8.120	7.717	312.8E6	299.4E6	563.539	594.762
9) T	2,4-D	8.349	8.011	375.6E6	396.0E6	654.067	660.909
10) T	Pentachlo...	8.662	8.442	3286.9E6	2853.2E6	442.334	457.239
11) T	2,4,5-TP ...	9.212	8.802	1856.4E6	1922.5E6	627.284	735.125m
12) T	2,4,5-T	9.497	9.182	1661.1E6	1574.1E6	599.094	603.216
13) T	2,4-DB	10.049	9.688	199.1E6	203.1E6	525.528	572.826
14) T	DINOSEB	11.212	10.024	788.3E6	731.2E6	352.289m	371.317
15) T	Picloram	11.027	11.002	1277.2E6	1233.6E6	412.265	362.705
16) T	DCPA	11.478	10.952	2157.8E6	2017.5E6	579.470m	601.692

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050120\
Data File : PS010107.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2020 19:56
Operator : DD\AJ
Sample : L2122-15MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SU-03-043020MS

Manual Integrations
APPROVED

Sohil
5/4/2020 8:10:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 03:44:54 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

