

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050120\
 Data File : PS010109.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2020 20:42
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 5/4/2020 8:10:48 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 03:45:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 09 05:27:48 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.922	486.1E6	401.5E6	805.001	810.719
Target Compounds							
1) T	Dalapon	3.045	2.626	468.0E6	439.1E6	617.716	613.295
2) T	3,5-DICHL...	6.446	6.071	614.2E6	555.9E6	746.887	748.266
3) T	4-Nitroph...	7.064	6.525	208.5E6	224.5E6	749.342	739.302
5) T	DICAMBA	7.427	7.091	1747.5E6	1521.7E6	775.799	765.064
6) T	MCPD	7.597	7.190	135.8E6	126.9E6	76.033	77.114
7) T	MCPA	7.746	7.405	194.0E6	181.5E6	75.247	76.083
8) T	DICHLORPROP	8.120	7.717	414.5E6	385.8E6	746.849	766.458
9) T	2,4-D	8.349	8.011	439.6E6	460.9E6	765.532	769.113
10) T	Pentachlo...	8.662	8.443	5791.3E6	4875.9E6	779.364	781.397
11) T	2,4,5-TP ...	9.212	8.803	2341.3E6	2062.0E6	791.149	788.477
12) T	2,4,5-T	9.495	9.182	2196.6E6	2035.1E6	792.247	779.875
13) T	2,4-DB	10.049	9.688	289.9E6	269.2E6	765.417	759.270
14) T	DINOSEB	11.213	10.023	1667.1E6	1506.8E6	745.028m	765.150m
15) T	Picloram	11.028	11.002	2107.5E6	2226.0E6	680.251	654.476
16) T	DCPA	11.478	10.952	2863.7E6	2617.8E6	769.048	780.726

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050120\
Data File : PS010109.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2020 20:42
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
5/4/2020 8:10:48 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 03:45:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040820.M
Quant Title : 8080.M
QLast Update : Thu Apr 09 05:27:48 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

