

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050421\
 Data File : PS014911.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2021 22:03
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 03:30:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.155	6.884	1233.6E6	349.7E6	810.929	751.341
Target Compounds							
1) T	Dalapon	3.005	2.602	1205.9E6	403.4E6	668.765	619.869
2) T	3,5-DICHL...	6.381	6.047	1581.0E6	519.3E6	763.030	708.131
3) T	4-Nitroph...	6.990	6.503	577.0E6	175.0E6	710.031	619.927
5) T	DICAMBA	7.342	7.050	4454.3E6	1476.9E6	765.564	730.247
6) T	MCPD	7.507	7.144	322.2E6	121.7E6	73.991	72.282
7) T	MCPA	7.656	7.359	464.3E6	170.1E6	74.615	70.024
8) T	DICHLORPROP	8.027	7.675	1061.4E6	360.7E6	758.489	714.572
9) T	2,4-D	8.258	7.972	1176.3E6	417.0E6	739.835	749.915
10) T	Pentachlo...	8.561	8.393	15975.3E6	4957.2E6	760.122	737.396
11) T	2,4,5-TP ...	9.106	8.755	6171.6E6	2001.8E6	753.877	732.822
12) T	2,4,5-T	9.392	9.135	6119.2E6	1889.1E6	743.877	722.250
13) T	2,4-DB	9.940	9.641	1098.7E6	247.3E6	722.009	712.800
14) T	DINOSEB	11.088	9.969	4837.5E6	1394.5E6	750.733	724.175
15) T	Picloram	10.917	10.956	7397.4E6	2431.4E6	661.177	648.389
16) T	DCPA	11.353	10.895	9162.7E6	2986.1E6	743.600	735.378

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050421\
Data File : PS014911.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2021 22:03
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 03:30:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

