

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050422\
 Data File : PS019266.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2022 16:33
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 17:09:34 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.220	6.799	2212.0E6	556.2E6	809.981	718.126
Target Compounds							
1) T	Dalapon	3.027	2.532	3262.5E6	599.3E6	675.789	674.216
2) T	3,5-DICHL...	6.431	5.975	2463.5E6	810.1E6	696.956	694.650
3) T	4-Nitroph...	7.055	6.417	745.7E6	240.6E6	769.529	681.776
5) T	DICAMBA	7.410	6.964	7825.4E6	2379.3E6	708.550	710.060
6) T	MCPD	7.581	7.062	643.3E6	188.1E6	71.961	70.723
7) T	MCPA	7.730	7.276	885.7E6	258.6E6	71.062	69.431
8) T	DICHLORPROP	8.104	7.587	1962.3E6	599.5E6	701.415	699.494
9) T	2,4-D	8.336	7.880	2247.2E6	699.0E6	712.316	699.975
10) T	Pentachlo...	8.643	8.306	28466.7E6	7961.5E6	724.591	708.227
11) T	2,4,5-TP ...	9.194	8.668	12860.0E6	3436.3E6	711.476	714.028
12) T	2,4,5-T	9.481	9.047	14917.7E6	3116.5E6	705.293	719.734
13) T	2,4-DB	10.036	9.553	3042.3E6	403.5E6	709.449	709.698
14) T	DINOSEB	11.190	9.878	11484.1E6	2052.6E6	695.530	708.312
15) T	Picloram	11.027	10.869	21607.1E6	4177.5E6	739.206	698.482
16) T	DCPA	11.458	10.809	18550.6E6	4198.8E6	708.824	715.240

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050422\
Data File : PS019266.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2022 16:33
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 17:09:34 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
Quant Title : 8080.M
QLast Update : Wed May 04 16:27:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

