

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050422\
 Data File : PS019271.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2022 18:34
 Operator : AJ\MA
 Sample : N2600-01MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-15SMSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/05/2022
 Supervised By : mohammad ahmed 05/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 19:33:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.220	6.797	1466.0E6	373.5E6	536.838m	482.255
Target Compounds							
1) T	Dalapon	3.026	2.532	853.4E6	184.0E6	176.780	207.014m
2) T	3,5-DICHL...	6.431	5.974	536.1E6	156.0E6	151.671	133.767
3) T	4-Nitroph...	7.051	6.414	73226335	1836411	75.562	5.205 #
5) T	DICAMBA	7.411	6.963	2212.9E6	642.3E6	200.364	191.679
6) T	MCPD	7.577	7.055	102.0E6	29901516	11.409	11.244
7) T	MCPA	7.725	7.267	196.6E6	75659600	15.773	20.313 #
8) T	DICHLORPROP	8.104	7.586	381.3E6	130.2E6	136.302	151.935
9) T	2,4-D	8.334	7.878	764.2E6	206.7E6	242.241	207.006
10) T	Pentachlo...	8.643	8.305	2949.3E6	816.9E6	75.072	72.664
11) T	2,4,5-TP ...	9.196	8.667	2111.1E6	552.7E6	116.798	114.838
12) T	2,4,5-T	9.484	9.050	2998.1E6	674.3E6	141.745	155.722
13) T	2,4-DB	10.035	9.564	253.3E6	324.4E6	59.070	570.615 #
14) T	DINOSEB	11.159	9.881	4934.3E6	577.9E6	298.842	199.406 #
15) T	Picloram	11.017	10.861	3563.4E6	1893.7E6	121.909m	316.622 #
16) T	DCPA	11.460	10.809	6367.9E6	1781.9E6	243.320	303.544

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050422\
 Data File : PS019271.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2022 18:34
 Operator : AJ\MA
 Sample : N2600-01MSD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 MH-15SMSD

Manual Integrations APPROVED

Reviewed By : Abdul Mirza 05/05/2022
 Supervised By : mohammad ahmed 05/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 19:33:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

