

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050520\
 Data File : PS010125.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2020 12:03
 Operator : DD\AJ
 Sample : HSTDICC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 13:40:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050520.M
 Quant Title : 8080.M
 QLast Update : Tue May 05 13:39:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.235	6.922	344.5E6	263.2E6	515.866	518.794
Target Compounds							
1) T	Dalapon	3.043	2.625	376.9E6	356.3E6	471.144	475.229
2) T	3,5-DICHL...	6.445	6.070	434.7E6	371.2E6	480.239	485.233
3) T	4-Nitroph...	7.064	6.526	147.3E6	152.1E6	478.507	482.451
5) T	DICAMBA	7.427	7.091	1251.4E6	1015.3E6	489.009	495.898
6) T	MCPD	7.595	7.186	96830892	81439503	47.560	48.097
7) T	MCPA	7.744	7.401	136.5E6	117.4E6	47.529	47.812
8) T	DICHLORPROP	8.121	7.718	298.8E6	253.2E6	484.812	488.956
9) T	2,4-D	8.350	8.012	305.7E6	304.3E6	464.793	488.838
10) T	Pentachlo...	8.662	8.443	4357.9E6	3366.7E6	512.852	538.559
11) T	2,4,5-TP ...	9.212	8.804	1700.0E6	1373.6E6	495.643	506.170
12) T	2,4,5-T	9.497	9.182	1617.2E6	1372.8E6	495.037	503.161
13) T	2,4-DB	10.050	9.689	215.2E6	181.1E6	470.604	492.810
14) T	DINOSEB	11.214	10.024	1217.9E6	999.5E6	491.985	497.077
15) T	Picloram	11.029	11.003	1896.3E6	1917.2E6	478.094	491.023
16) T	DCPA	11.479	10.953	2166.2E6	1759.7E6	508.062	514.422

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050520\
Data File : PS010125.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2020 12:03
Operator : DD\AJ
Sample : HSTDICC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 13:40:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050520.M
Quant Title : 8080.M
QLast Update : Tue May 05 13:39:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

