

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050520\
 Data File : PS010126.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2020 12:26
 Operator : DD\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 13:40:14 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050520.M
 Quant Title : 8080.M
 QLast Update : Tue May 05 13:39:05 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.921	500.8E6	380.6E6	750.000	750.000
Target Compounds							
1) T	Dalapon	3.042	2.623	546.0E6	511.7E6	682.500	682.500
2) T	3,5-DICHL...	6.446	6.070	631.4E6	533.5E6	697.500	697.500
3) T	4-Nitroph...	7.064	6.525	210.1E6	215.2E6	682.500	682.500
5) T	DICAMBA	7.427	7.091	1804.2E6	1443.4E6	705.000	705.000
6) T	MCPD	7.597	7.188	143.5E6	119.4E6	70.500	70.500
7) T	MCPA	7.747	7.404	200.3E6	171.3E6	69.750	69.750
8) T	DICHLORPROP	8.120	7.717	434.5E6	365.1E6	705.000	705.000
9) T	2,4-D	8.349	8.011	463.6E6	438.8E6	705.000	705.000
10) T	Pentachlo...	8.662	8.443	6054.4E6	4454.1E6	712.500	712.500
11) T	2,4,5-TP ...	9.212	8.803	2443.8E6	1933.6E6	712.500	712.500
12) T	2,4,5-T	9.496	9.182	2327.6E6	1944.0E6	712.500	712.500
13) T	2,4-DB	10.049	9.688	325.8E6	261.8E6	712.500	712.500
14) T	DINOSEB	11.214	10.023	1745.2E6	1417.6E6	705.000	705.000
15) T	Picloram	11.029	11.003	2826.0E6	2782.0E6	712.500	712.500
16) T	DCPA	11.478	10.953	3069.8E6	2462.9E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050520\
Data File : PS010126.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2020 12:26
Operator : DD\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 13:40:14 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050520.M
Quant Title : 8080.M
QLast Update : Tue May 05 13:39:05 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

