

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050522\
 Data File : PS019333.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2022 02:46
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 03:02:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.219	6.798	2182.3E6	547.8E6	799.115	707.210
Target Compounds							
1) T	Dalapon	3.027	2.533	3133.9E6	579.4E6	649.144	651.836
2) T	3,5-DICHL...	6.430	5.975	2464.5E6	798.3E6	697.240	684.471
3) T	4-Nitroph...	7.053	6.416	802.6E6	246.8E6	828.231	699.575
5) T	DICAMBA	7.409	6.963	7765.2E6	2332.8E6	703.093	696.200
6) T	MCPD	7.580	7.061	634.2E6	183.8E6	70.941	69.132
7) T	MCPA	7.729	7.275	873.4E6	254.4E6	70.075	68.311
8) T	DICHLORPROP	8.103	7.586	1937.4E6	591.6E6	692.541	690.277
9) T	2,4-D	8.334	7.879	2259.3E6	694.6E6	716.144	695.572
10) T	Pentachlo...	8.642	8.305	28024.5E6	7880.3E6	713.335	701.007
11) T	2,4,5-TP ...	9.194	8.667	12553.5E6	3389.4E6	694.516	704.284
12) T	2,4,5-T	9.480	9.045	14396.9E6	3114.7E6	680.671	719.329
13) T	2,4-DB	10.035	9.551	3021.9E6	400.5E6	704.682	704.456
14) T	DINOSEB	11.188	9.877	11732.3E6	2175.7E6	710.561	750.800
15) T	Picloram	11.023	10.866	20427.7E6	3983.7E6	698.857	666.089
16) T	DCPA	11.456	10.807	18242.5E6	4151.5E6	697.050	707.185

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050522\
Data File : PS019333.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2022 02:46
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 03:02:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
Quant Title : 8080.M
QLast Update : Wed May 04 16:27:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

