

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050621\
 Data File : PS014930.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2021 20:33
 Operator : DD\AJ
 Sample : M2280-01MSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-PS-01-072MSD

Manual Integrations
 APPROVED

Ankita
 5/7/2021 4:18:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 04:07:07 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.153	6.883	419.5E6	132.9E6	275.746	285.495m
Target Compounds							
1) T	Dalapon	3.003	2.604	825.5E6	369.6E6	457.812m	567.871
2) T	3,5-DICHL...	6.380	6.047	643.1E6	191.3E6	310.357	260.833
5) T	DICAMBA	7.340	7.048	2206.1E6	744.1E6	379.155	367.918
6) T	MCPD	7.503	7.139	102.9E6	41549342	23.635	24.672
7) T	MCPA	7.651	7.352	190.7E6	112.6E6	30.649	46.355 #
8) T	DICHLORPROP	8.027	7.673	389.9E6	157.3E6	278.603	311.515
9) T	2,4-D	8.256	7.970	665.8E6	235.1E6	418.744	422.859
10) T	Pentachlo...	8.561	8.393	4532.0E6	1468.4E6	215.638	218.431
11) T	2,4,5-TP ...	9.104	8.753	1867.9E6	641.7E6	228.165	234.893
12) T	2,4,5-T	9.390	9.133	2254.2E6	720.3E6	274.025	275.382
13) T	2,4-DB	9.939	9.641	250.4E6	90850226	164.537	261.913 #
14) T	DINOSEB	11.084	9.985	249.2E6	59127939	38.676	30.705
15) T	Picloram	10.915	10.953	4765.7E6	1641.0E6	425.954	437.614
16) T	DCPA	11.354	10.894	4726.1E6	1704.3E6	383.548	419.711

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050621\
Data File : PS014930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2021 20:33
Operator : DD\AJ
Sample : M2280-01MSD
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
FK-PS-01-072MSD

Manual Integrations
APPROVED

Ankita
5/7/2021 4:18:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 04:07:07 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

