

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050622\
 Data File : PS019338.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2022 12:28
 Operator : AJ\MA
 Sample : N2548-10MSD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-11MSD

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/09/2022
 Supervised By : mohammad ahmed 05/09/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 22:20:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.221	6.797	1836.4E6	543.4E6	672.475	701.626
Target Compounds							
1) T	Dalapon	3.026	2.530	1230.3E6	236.1E6	254.850	265.608
2) T	3,5-DICHL...	6.432	5.973	895.6E6	297.2E6	253.384	254.803
3) T	4-Nitroph...	7.051	6.413	359.1E6	94355182	370.540	267.421 #
5) T	DICAMBA	7.412	6.962	3000.0E6	881.2E6	271.629	262.977
6) T	MCPD	7.579	7.056	202.3E6	54505290	22.625	20.496
7) T	MCPA	7.727	7.268	286.7E6	97059296	23.005	26.059
8) T	DICHLORPROP	8.107	7.585	633.7E6	213.2E6	226.501	248.829
9) T	2,4-D	8.337	7.878	950.7E6	303.7E6	301.349	304.070
10) T	Pentachlo...	8.646	8.305	8380.9E6	2354.1E6	213.326	209.416
11) T	2,4,5-TP ...	9.198	8.667	3893.3E6	1043.3E6	215.394	216.792
12) T	2,4,5-T	9.486	9.049	4729.8E6	1051.4E6	223.618	242.818
13) T	2,4-DB	10.037	9.561	567.4E6	320.1E6	132.307	562.952 #
14) T	DINOSEB	11.160	9.880	3883.7E6	561.6E6	235.213m	193.797
15) T	Picloram	11.018	10.861	9525.7E6	2202.6E6	325.886m	368.285
16) T	DCPA	11.463	10.808	7407.1E6	1999.5E6	283.027	340.598

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050622\
Data File : PS019338.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2022 12:28
Operator : AJ\MA
Sample : N2548-10MSD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-11MSD

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/09/2022
Supervised By : mohammad ahmed 05/09/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 22:20:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
Quant Title : 8080.M
QLast Update : Wed May 04 16:27:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

