

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050624\
 Data File : PS026659.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2024 17:37
 Operator : AR\AJ
 Sample : P2323-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 TP-3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 01:08:54 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050624.M
 Quant Title : 8080.M
 QLast Update : Mon May 06 13:55:47 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.276	7.783	871.8E6	497.8E6	350.558	298.527
Target Compounds							
1) T	Dalapon	0.000	2.678f	0	528.2E6	N.D.	180.157
2) T	3,5-DICHL...	6.426	6.739	4966918	5223593	1.444	2.135 #
3) T	4-Nitroph...	7.070	7.296	13439389	15339038	8.341	13.488 #
5) T	DICAMBA	7.485	7.980	21420312	-2402199	2.089	N.D. #
7) T	MCPA	7.836f	8.325	52873524	28918589	4.777	3.867
8) T	DICHLORPROP	8.138f	8.678	30939429	7313436	10.324	3.812 #
9) T	2,4-D	8.418	9.026	8841214	5230969	2.686	2.539
10) T	Pentachlo...	8.696	9.560	15495688	6622191	<MDL	<MDL #
11) T	2,4,5-TP ...	9.235f	9.906f	8295698	18999.6E6	<MDL	1721.742 #
12) T	2,4,5-T	0.000	10.379	0	7358199	N.D.	<MDL
14) T	DINOSEB	0.000	11.304	0	50149544	N.D.	7.301
15) T	Picloram	11.137f	12.413	-70049890	219.8E6	N.D.	13.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050624\
Data File : PS026659.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2024 17:37
Operator : AR\AJ
Sample : P2323-07
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
TP-3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 01:08:54 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050624.M
Quant Title : 8080.M
QLast Update : Mon May 06 13:55:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 2.663 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
TP-3




#1 Dalapon

R.T.: 2.678 min
Delta R.T.: -0.035 min
Response: 528176462
Conc: 180.16 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.426 min
Delta R.T.: -0.017 min
Response: 4966918
Conc: 1.44 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.739 min
Delta R.T.: 0.009 min
Response: 5223593
Conc: 2.14 ng/ml



#3 4-Nitrophenol


R.T.: 7.070 min
Delta R.T.: -0.009 min
Response: 13439389
Conc: 8.34 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3




#3 4-Nitrophenol

R.T.: 7.296 min
Delta R.T.: -0.013 min
Response: 15339038
Conc: 13.49 ng/ml




#4 2,4-DCAA

R.T.: 7.276 min
Delta R.T.: -0.001 min
Response: 871789010
Conc: 350.56 ng/ml



#4 2,4-DCAA


R.T.: 7.783 min
Delta R.T.: -0.001 min
Response: 497768874
Conc: 298.53 ng/ml



#5 DICAMBA


R.T.: 7.485 min
Delta R.T.: 0.020 min
Response: 21420312
Conc: 2.09 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3




#5 DICAMBA

R.T.: 7.980 min
Delta R.T.: -0.004 min
Response: -2402199
Conc: N.D.




#7 MCPA

R.T.: 7.836 min
Delta R.T.: 0.044 min
Response: 52873524
Conc: 4.78 ug/ml



#7 MCPA


R.T.: 8.325 min
Delta R.T.: -0.001 min
Response: 28918589
Conc: 3.87 ug/ml



#8 DICHLORPROP


R.T.: 8.138 min
Delta R.T.: -0.033 min
Response: 30939429
Conc: 10.32 ng/ml

Instrument :
ECD_S
ClientSampleId :
TP-3




#8 DICHLORPROP

R.T.: 8.678 min
Delta R.T.: -0.023 min
Response: 7313436
Conc: 3.81 ng/ml




#9 2,4-D

R.T.: 8.418 min
Delta R.T.: 0.015 min
Response: 8841214
Conc: 2.69 ng/ml



#9 2,4-D


R.T.: 9.026 min
Delta R.T.: -0.006 min
Response: 5230969
Conc: 2.54 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.235 min
Delta R.T.: -0.046 min
Response: 8295698
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
TP-3




#11 2,4,5-TP (SILVEX)

R.T.: 9.906 min
Delta R.T.: -0.032 min
Response: 18999645711
Conc: 1721.74 ng/ml




#13 2,4-DB

R.T.: 10.182 min
Delta R.T.: 0.036 min
Response: 1235681019
Conc: 521.19 ng/ml



#13 2,4-DB


R.T.: 10.953 min
Delta R.T.: 0.029 min
Response: -991568525
Conc: N.D.



#14 DINOSEB


R.T.: 0.000 min
Exp R.T.: 11.361 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
TP-3




#14 DINOSEB

R.T.: 11.304 min
Delta R.T.: 0.001 min
Response: 50149544
Conc: 7.30 ng/ml



#15 Picloram

R.T.: 11.137 min
Delta R.T.: -0.031 min
Response: -70049890
Conc: N.D.



#15 Picloram

R.T.: 12.413 min
Delta R.T.: 0.012 min
Response: 219823972
Conc: 13.72 ng/ml