

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050721\
 Data File : PS014948.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2021 19:26
 Operator : DD\AJ
 Sample : PB136255TB
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB136255TB

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 05:51:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
 Quant Title : 8080.M
 QLast Update : Thu Apr 08 14:34:06 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.156	6.885	616.3E6	209.7E6	405.121	450.539
Target Compounds							
1) T	Dalapon	2.962f	2.568	103.2E6	142.5E6	57.216	219.006 #
2) T	3,5-DICHL...	0.000	6.075	0	62653753	N.D.	85.441
3) T	4-Nitroph...	6.990	6.469	6320856	16172003	7.778	57.280 #
5) T	DICAMBA	7.330	7.018	19255565	12781062	3.309	6.319 #
6) T	MCPD	7.502	7.094f	3506521	6559171	<MDL	3.895 #
7) T	MCPA	7.692	0.000	5609409	0	<MDL	N.D. #
8) T	DICHLORPROP	8.041	7.723f	4948656	13596175	3.536	26.933 #
9) T	2,4-D	8.322f	0.000	18257100	0	11.483	N.D. #
10) T	Pentachlo...	8.565	8.397	9609289	4885446	<MDL	<MDL #
11) T	2,4,5-TP ...	9.109	8.731	5931467	13028290	<MDL	4.769 #
13) T	2,4-DB	9.979	0.000	10188803	0	6.695	N.D. #
14) T	DINOSEB	11.067	10.010	39709684	2762400	6.163	1.435 #
15) T	Picloram	10.959f	10.990	32643446	8888054	2.918	2.370
16) T	DCPA	11.358	10.897	6754557	4516904	<MDL	1.112 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050721\
Data File : PS014948.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2021 19:26
Operator : DD\AJ
Sample : PB136255TB
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB136255TB

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 05:51:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040821.M
Quant Title : 8080.M
QLast Update : Thu Apr 08 14:34:06 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.962 min
Delta R.T.: -0.044 min
Response: 103171146
Conc: 57.22 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB136255TB




#1 Dalapon

R.T.: 2.568 min
Delta R.T.: -0.035 min
Response: 142534656
Conc: 219.01 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 6.384 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 6.075 min
Delta R.T.: 0.030 min
Response: 62653753
Conc: 85.44 ng/ml



#3 4-Nitrophenol


R.T.: 6.990 min
Delta R.T.: 0.000 min
Response: 6320856
Conc: 7.78 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB136255TB




#3 4-Nitrophenol

R.T.: 6.469 min
Delta R.T.: -0.028 min
Response: 16172003
Conc: 57.28 ng/ml




#4 2,4-DCAA

R.T.: 7.156 min
Delta R.T.: -0.002 min
Response: 616267460
Conc: 405.12 ng/ml



#4 2,4-DCAA


R.T.: 6.885 min
Delta R.T.: 0.002 min
Response: 209705570
Conc: 450.54 ng/ml



#5 DICAMBA


R.T.: 7.330 min
Delta R.T.: -0.017 min
Response: 19255565
Conc: 3.31 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB136255TB




#5 DICAMBA

R.T.: 7.018 min
Delta R.T.: -0.032 min
Response: 12781062
Conc: 6.32 ng/ml




#6 MCPP

R.T.: 7.502 min
Delta R.T.: -0.018 min
Response: 3506521
Conc: N.D.



#6 MCPP


R.T.: 7.094 min
Delta R.T.: -0.059 min
Response: 6559171
Conc: 3.89 ug/ml



#8 DICHLORPROP


R.T.: 8.041 min
Delta R.T.: 0.008 min
Response: 4948656
Conc: 3.54 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB136255TB




#8 DICHLORPROP

R.T.: 7.723 min
Delta R.T.: 0.048 min
Response: 13596175
Conc: 26.93 ng/ml




#9 2,4-D

R.T.: 8.322 min
Delta R.T.: 0.061 min
Response: 18257100
Conc: 11.48 ng/ml



#9 2,4-D


R.T.: 0.000 min
Exp R.T. : 7.970 min
Response: 0
Conc: N.D.



#11 2,4,5-TP (SILVEX)


R.T.: 9.109 min
Delta R.T.: -0.003 min
Response: 5931467
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
PB136255TB




#11 2,4,5-TP (SILVEX)

R.T.: 8.731 min
Delta R.T.: -0.025 min
Response: 13028290
Conc: 4.77 ng/ml




#13 2,4-DB

R.T.: 9.979 min
Delta R.T.: 0.036 min
Response: 10188803
Conc: 6.70 ng/ml



#13 2,4-DB


R.T.: 0.000 min
Exp R.T.: 9.640 min
Response: 0
Conc: N.D.



#14 DINOSEB


R.T.: 11.067 min
Delta R.T.: -0.029 min
Response: 39709684
Conc: 6.16 ng/ml

Instrument :
ECD_S
ClientSampleId :
PB136255TB




#14 DINOSEB

R.T.: 10.010 min
Delta R.T.: 0.038 min
Response: 2762400
Conc: 1.43 ng/ml




#15 Picloram

R.T.: 10.959 min
Delta R.T.: 0.040 min
Response: 32643446
Conc: 2.92 ng/ml



#15 Picloram


R.T.: 10.990 min
Delta R.T.: 0.038 min
Response: 8888054
Conc: 2.37 ng/ml



#16 DCPA

R.T.: 11.358 min
Delta R.T.: -0.005 min
Response: 6754557
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
PB136255TB



#16 DCPA

R.T.: 10.897 min
Delta R.T.: 0.000 min
Response: 4516904
Conc: 1.11 ng/ml