

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050820\
 Data File : PS010167.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 May 2020 14:12
 Operator : DD\AJ
 Sample : PB128704BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB128704BS

Manual Integrations
 APPROVED

Sohil
 5/11/2020 9:07:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 16:20:36 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050520.M
 Quant Title : 8080.M
 QLast Update : Tue May 05 13:43:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.234	6.921	343.1E6	285.9E6	507.245	554.447
Target Compounds							
1) T	Dalapon	3.041	2.623	319.4E6	418.3E6	392.111	552.134m#
2) T	3,5-DICHL...	6.445	6.070	442.3E6	354.3E6	485.113	458.920
3) T	4-Nitroph...	7.064	6.525	133.1E6	130.7E6	421.324	406.855
5) T	DICAMBA	7.426	7.090	1174.4E6	963.4E6	459.906	469.928
6) T	MCPD	7.594	7.186	111.6E6	81647219	54.413	47.757
7) T	MCPA	7.743	7.400	134.7E6	141.8E6	46.256	56.938
8) T	DICHLORPROP	8.120	7.717	307.1E6	266.6E6	491.055	506.487
9) T	2,4-D	8.348	8.010	359.7E6	351.5E6	546.573	555.428
10) T	Pentachlo...	8.661	8.442	4345.8E6	3402.7E6	509.142	531.101
11) T	2,4,5-TP ...	9.211	8.803	1780.7E6	1458.0E6	518.598	534.945
12) T	2,4,5-T	9.499	9.182	1580.6E6	1372.2E6	483.738	503.198
13) T	2,4-DB	10.049	9.688	189.3E6	180.2E6	415.634	480.451
14) T	DINOSEB	11.212	10.023	1178.5E6	995.5E6	471.325m	488.202
15) T	Picloram	11.027	11.002	1890.6E6	1649.1E6	485.117	432.330
16) T	DCPA	11.478	10.952	2006.9E6	1782.6E6	468.432	518.572

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050820\
Data File : PS010167.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 May 2020 14:12
Operator : DD\AJ
Sample : PB128704BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB128704BS

Manual Integrations
APPROVED

Sohil
5/11/2020 9:07:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 16:20:36 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050520.M
Quant Title : 8080.M
QLast Update : Tue May 05 13:43:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

