

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050919\
 Data File : PS004523.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2019 21:04
 Operator : EI\SJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 5/10/2019 2:48:52 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 01:06:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS042919.M
 Quant Title : 8080.M
 QLast Update : Tue Apr 30 08:27:54 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.012	5.709	487.1E6	374.6E6	773.920m	776.774
Target Compounds							
1) T	Dalapon	2.408	2.066	560.5E6	430.1E6	713.268	642.757
2) T	3,5-DICHL...	5.352	4.996	602.4E6	422.3E6	762.526m	686.922
3) T	4-Nitroph...	5.860	5.382	199.3E6	169.1E6	797.430	683.617
5) T	DICAMBA	6.174	5.843	1893.3E6	1335.2E6	796.167	713.367
6) T	MCPD	6.331	5.929	134.3E6	91320960	77.718	67.906
7) T	MCPA	6.468	6.114	193.0E6	130.5E6	74.917	65.708
8) T	DICHLORPROP	6.817	6.402	446.6E6	324.1E6	780.310m	688.770
9) T	2,4-D	7.035	6.675	454.3E6	363.4E6	798.437m	691.124
10) T	Pentachlo...	7.305	7.052	6713.5E6	4749.1E6	788.043	712.024
11) T	2,4,5-TP ...	7.844	7.416	2626.2E6	1849.4E6	814.379	703.868
12) T	2,4,5-T	8.120	7.814	2674.3E6	1551.0E6	785.479	689.576
13) T	2,4-DB	8.654	8.267	319.6E6	206.7E6	949.880	753.847
14) T	DINOSEB	9.756	8.582	2126.0E6	1316.1E6	707.724	673.348
15) T	Picloram	9.585	9.524	3400.3E6	2524.7E6	803.279	690.318
16) T	DCPA	10.026	9.483	3419.9E6	2366.8E6	782.730m	719.527

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS050919\
Data File : PS004523.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2019 21:04
Operator : EI\SJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
5/10/2019 2:48:52 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 01:06:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS042919.M
Quant Title : 8080.M
QLast Update : Tue Apr 30 08:27:54 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

