

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051019\
 Data File : PS004528.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2019 16:04
 Operator : EI\SJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Ankita
 5/13/2019 11:03:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 03:32:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051019.M
 Quant Title : 8080.M
 QLast Update : Sat May 11 03:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.027	5.712	122.4E6	94499209	196.605	198.426
Target Compounds							
1) T	Dalapon	2.405	2.063	155.7E6	111.9E6	167.773	175.990
2) T	3,5-DICHL...	5.364	4.999	152.3E6	118.7E6	180.967	187.658
3) T	4-Nitroph...	5.934	5.395	27429019	33066386	119.712m	153.967 #
5) T	DICAMBA	6.178	5.842	467.6E6	376.8E6	183.535	190.500
6) T	MCPD	6.328	5.923	28772573	21044810	15.318	15.429
7) T	MCPA	6.465	6.108	43903802	33473194	16.158	16.965
8) T	DICHLORPROP	6.830	6.405	115.2E6	92166453	186.451	191.982
9) T	2,4-D	7.078	6.685	107.2E6	96737461	179.696m	183.687
10) T	Pentachlo...	7.307	7.050	1878.0E6	1424.0E6	199.623	203.880
11) T	2,4,5-TP ...	7.853	7.417	656.6E6	509.2E6	184.867	189.099
12) T	2,4,5-T	8.146	7.820	650.9E6	467.2E6	181.809	197.330
13) T	2,4-DB	8.691	8.275	44523574	49031853	130.430	171.868 #
14) T	DINOSEB	9.758	8.581	700.6E6	374.6E6	216.140	186.540
15) T	Picloram	9.622	9.532	579.2E6	618.4E6	130.876m	169.005 #
16) T	DCPA	10.030	9.481	905.8E6	686.0E6	195.911	200.019

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051019\
 Data File : PS004528.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2019 16:04
 Operator : EI\SJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Ankita
 5/13/2019 11:03:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 03:32:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051019.M
 Quant Title : 8080.M
 QLast Update : Sat May 11 03:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

