

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051019\
 Data File : PS004532.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2019 18:06
 Operator : EI\SJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Manual Integrations
 APPROVED

Ankita
 5/13/2019 11:03:04 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 04:10:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051019.M
 Quant Title : 8080.M
 QLast Update : Sat May 11 03:30:59 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.014	5.708	929.6E6	703.6E6	1493.015	1477.337
Target Compounds							
1) T	Dalapon	2.406	2.063	1211.6E6	826.8E6	1305.777	1300.373
2) T	3,5-DICHL...	5.353	4.994	1153.9E6	866.5E6	1371.319	1369.808
3) T	4-Nitroph...	5.862	5.380	337.1E6	293.5E6	1471.305	1366.572
5) T	DICAMBA	6.174	5.842	3515.1E6	2601.1E6	1379.641	1315.141
6) T	MCP	6.334	5.933	224.5E6	190.1E6	119.524	139.405m
7) T	MCPA	6.473	6.119	388.3E6	274.9E6	142.906m	139.311
8) T	DICHLORPROP	6.820	6.402	860.8E6	608.0E6	1393.084	1266.534
9) T	2,4-D	7.038	6.675	859.0E6	725.8E6	1439.798m	1378.187
10) T	Pentachlo...	7.305	7.051	11950.1E6	8733.9E6	1270.237	1250.517
11) T	2,4,5-TP ...	7.845	7.415	4800.2E6	3590.5E6	1351.437	1333.341
12) T	2,4,5-T	8.122	7.813	4833.1E6	3163.3E6	1350.008	1335.966
13) T	2,4-DB	8.656	8.266	532.0E6	444.4E6	1558.348m	1557.769
14) T	DINOSEB	9.756	8.581	4074.2E6	2748.5E6	1256.957	1368.690
15) T	Picloram	9.587	9.523	6339.2E6	5101.6E6	1432.500	1394.230
16) T	DCPA	10.028	9.481	5958.4E6	4540.8E6	1288.644	1324.051

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051019\
Data File : PS004532.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2019 18:06
Operator : EI\SJ
Sample : HSTDIC1500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDIC1500

Manual Integrations
APPROVED

Ankita
5/13/2019 11:03:04 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 04:10:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051019.M
Quant Title : 8080.M
QLast Update : Sat May 11 03:30:59 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

