

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051022\
 Data File : PS019381.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2022 17:46
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 01:14:12 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
 Quant Title : 8080.M
 QLast Update : Wed May 04 16:27:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.220	6.798	2190.8E6	587.8E6	802.242	758.825
Target Compounds							
1) T	Dalapon	3.025	2.531	3278.7E6	647.9E6	679.150	728.824
2) T	3,5-DICHL...	6.431	5.975	2490.4E6	849.6E6	704.560	728.457
3) T	4-Nitroph...	7.053	6.416	863.2E6	274.3E6	890.687	777.487
5) T	DICAMBA	7.410	6.964	7795.3E6	2486.5E6	705.817	742.046
6) T	MCPD	7.581	7.062	641.4E6	197.2E6	71.753	74.138
7) T	MCPA	7.730	7.276	887.4E6	274.4E6	71.205	73.677
8) T	DICHLORPROP	8.103	7.586	1950.4E6	631.4E6	697.182	736.757
9) T	2,4-D	8.335	7.879	2305.5E6	753.4E6	730.783	754.395
10) T	Pentachlo...	8.642	8.306	27588.9E6	8299.5E6	702.246	738.297
11) T	2,4,5-TP ...	9.194	8.667	12357.5E6	3626.3E6	683.675	753.510
12) T	2,4,5-T	9.479	9.045	14216.5E6	3359.2E6	672.142	775.797
13) T	2,4-DB	10.033	9.552	3021.5E6	434.0E6	704.604	763.407
14) T	DINOSEB	11.188	9.877	11369.3E6	2380.2E6	688.575	821.369
15) T	Picloram	11.020	10.864	21456.7E6	4502.1E6	734.062	752.768
16) T	DCPA	11.455	10.807	17688.6E6	4405.3E6	675.884	750.423

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051022\
Data File : PS019381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2022 17:46
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 01:14:12 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS050422.M
Quant Title : 8080.M
QLast Update : Wed May 04 16:27:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

