

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051023\
 Data File : PS022959.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2023 14:58
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 15:50:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
 Quant Title : 8080.M
 QLast Update : Wed May 10 15:37:07 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.482	6.858	1112.0E6	414.0E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.550	2.580	1092.6E6	580.1E6	682.500	682.500
2) T	3,5-DICHL...	5.835	6.024	1374.7E6	587.2E6	697.500	697.500
3) T	4-Nitroph...	6.314	6.463	697.9E6	221.3E6	682.500	682.500
5) T	DICAMBA	6.631	7.025	4616.7E6	1764.4E6	705.000	705.000
6) T	MCP	6.790	7.120	339.3E6	126.5E6	70.500	70.500
7) T	MCPA	6.918	7.330	491.0E6	181.3E6	69.750	69.750
8) T	DICHLORPROP	7.242	7.647	1291.1E6	428.5E6	705.000	705.000
9) T	2,4-D	7.444	7.935	1601.4E6	501.1E6	705.000	705.000
10) T	Pentachlo...	7.678	8.367	18987.5E6	6664.2E6	712.500	712.500
11) T	2,4,5-TP ...	8.215	8.726	8228.1E6	2551.1E6	712.500	712.500
12) T	2,4,5-T	8.476	9.101	8142.0E6	2303.3E6	712.500	712.500
13) T	2,4-DB	8.991	9.607	1618.4E6	276.1E6	712.500	712.500
14) T	DINOSEB	10.067	9.936	5616.7E6	1661.5E6	705.000	705.000
15) T	Picloram	9.886	10.903	12115.5E6	3053.7E6	712.500	712.500
16) T	DCPA	10.336	10.866	9220.3E6	2697.2E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051023\
Data File : PS022959.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2023 14:58
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 15:50:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
Quant Title : 8080.M
QLast Update : Wed May 10 15:37:07 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

