

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051123\
 Data File : PS022992.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2023 14:34
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 21:16:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
 Quant Title : 8080.M
 QLast Update : Wed May 10 16:09:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.483	6.858	1097.1E6	416.8E6	721.274	742.755
Target Compounds							
1) T	Dalapon	2.550	2.580	1080.4E6	572.8E6	668.198	663.607
2) T	3,5-DICHL...	5.835	6.023	1378.1E6	589.4E6	686.728	694.884
3) T	4-Nitroph...	6.315	6.463	698.1E6	225.0E6	663.107	678.039
5) T	DICAMBA	6.632	7.025	4592.0E6	1771.6E6	699.300	711.073
6) T	MCP	6.791	7.120	314.6E6	125.8E6	67.332	71.122
7) T	MCPA	6.919	7.330	464.9E6	180.3E6	66.411	68.107
8) T	DICHLORPROP	7.243	7.647	1256.5E6	431.9E6	670.443	697.959
9) T	2,4-D	7.444	7.935	1563.6E6	505.7E6	675.635	706.818
10) T	Pentachlo...	7.679	8.367	18725.8E6	6716.2E6	699.512	724.544
11) T	2,4,5-TP ...	8.216	8.726	8046.0E6	2571.4E6	688.810	722.272
12) T	2,4,5-T	8.476	9.100	8020.6E6	2339.7E6	693.993	728.704
13) T	2,4-DB	8.991	9.607	1590.5E6	282.4E6	677.973	735.402
14) T	DINOSEB	10.066	9.936	5802.5E6	1732.0E6	710.761	732.177
15) T	Picloram	9.886	10.903	11929.3E6	3007.1E6	695.650	706.076
16) T	DCPA	10.335	10.865	9331.2E6	2729.1E6	715.297	730.357

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051123\
Data File : PS022992.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2023 14:34
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 21:16:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
Quant Title : 8080.M
QLast Update : Wed May 10 16:09:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

