

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051319\
 Data File : PS004573.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2019 13:42
 Operator : EI\SJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 15:14:19 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051319.M
 Quant Title : 8080.M
 QLast Update : Mon May 13 15:13:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.021	5.713	509.1E6	363.0E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.409	2.066	622.3E6	472.6E6	682.500	682.500
2) T	3,5-DICHL...	5.361	5.000	577.7E6	443.3E6	697.500	697.500
3) T	4-Nitroph...	5.876	5.387	176.6E6	146.1E6	682.500	682.500
5) T	DICAMBA	6.179	5.846	1849.6E6	1356.0E6	705.000	705.000
6) T	MCPD	6.335	5.932	137.3E6	98069776	70.500	70.500
7) T	MCPA	6.472	6.117	194.1E6	138.8E6	69.750	69.750
8) T	DICHLORPROP	6.826	6.406	427.0E6	329.8E6	705.000	705.000
9) T	2,4-D	7.049	6.681	554.2E6	368.6E6	705.000	705.000
10) T	Pentachlo...	7.311	7.055	6940.4E6	4864.1E6	712.500	712.500
11) T	2,4,5-TP ...	7.852	7.420	2697.2E6	1873.4E6	712.500	712.500
12) T	2,4,5-T	8.131	7.819	2721.8E6	1674.2E6	712.500	712.500
13) T	2,4-DB	8.666	8.272	252.6E6	218.1E6	712.500	712.500
14) T	DINOSEB	9.761	8.584	2370.5E6	1424.5E6	705.000	705.000
15) T	Picloram	9.597	9.528	3397.1E6	2599.0E6	712.500	712.500
16) T	DCPA	10.033	9.485	3497.2E6	2403.4E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051319\
Data File : PS004573.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2019 13:42
Operator : EI\SJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 15:14:19 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051319.M
Quant Title : 8080.M
QLast Update : Mon May 13 15:13:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

