

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051419\
 Data File : PS004593.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 May 2019 11:32
 Operator : EI\SJ
 Sample : K2774-01MS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-PS-01-048MS

Manual Integrations
 APPROVED

Ankita
 5/15/2019 1:57:15 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 01:05:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051319.M
 Quant Title : 8080.M
 QLast Update : Mon May 13 15:36:52 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.995	5.691	153.5E6	106.2E6	215.983	211.174
Target Compounds							
1) T	Dalapon	2.386	2.044	221.9E6	178.0E6	234.833m	253.858m
2) T	3,5-DICHL...	5.337	4.979	171.3E6	131.0E6	202.264	199.100m
3) T	4-Nitroph...	5.905	5.382	17614977	9728995	69.363m	44.003m#
5) T	DICAMBA	6.155	5.824	549.5E6	380.3E6	205.592	194.012m
6) T	MCPD	6.311	5.909	36943713	26740656	19.022m	19.179
7) T	MCPA	6.448	6.092	56551211	51424836	20.130m	25.162 #
8) T	DICHLORPROP	6.802	6.385	131.1E6	108.6E6	212.636	224.116
9) T	2,4-D	7.022	6.658	183.5E6	133.3E6	251.794m	247.109
10) T	Pentachlo...	7.288	7.034	1969.0E6	1509.2E6	206.243	216.938
11) T	2,4,5-TP ...	7.827	7.397	724.2E6	568.1E6	198.547	212.085
12) T	2,4,5-T	8.105	7.798	700.5E6	282.6E6	189.300	117.432 #
13) T	2,4-DB	8.641	8.250	95308378	60137897	264.395	202.974
14) T	DINOSEB	9.740	8.563	129.7E6	120.7E6	38.594	59.039 #
15) T	Picloram	9.575	9.508	1011.1E6	750.5E6	227.730	204.065
16) T	DCPA	10.010	9.465	741.6E6	413.9E6	156.913	121.362

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051419\
Data File : PS004593.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 May 2019 11:32
Operator : EI\SJ
Sample : K2774-01MS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleID :
FK-PS-01-048MS

Manual Integrations
APPROVED

Ankita
5/15/2019 1:57:15 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 01:05:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051319.M
Quant Title : 8080.M
QLast Update : Mon May 13 15:36:52 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

