

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051421\
 Data File : PS015018.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 May 2021 18:08
 Operator : DD\AJ
 Sample : M2364-02MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 AST2-COPR4-1MSD

Manual Integrations
 APPROVED

Ankita
 5/17/2021 5:12:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 18:46:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051021.M
 Quant Title : 8080.M
 QLast Update : Tue May 11 00:04:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.156	6.884	861.0E6	288.7E6	514.555	596.502m
Target Compounds							
1) T	Dalapon	3.002	2.602	822.1E6	382.5E6	401.923	565.914 #
2) T	3,5-DICHL...	6.381	6.047	1159.8E6	354.4E6	494.836	458.732
5) T	DICAMBA	7.342	7.049	3162.0E6	1082.3E6	482.562	505.337
6) T	MCP	7.507	7.142	241.9E6	90369510	49.469	51.130
7) T	MCPA	7.654	7.355	304.8E6	119.5E6	43.881m	47.593m
8) T	DICHLORPROP	8.028	7.674	810.9E6	296.7E6	508.415	548.591
9) T	2,4-D	8.259	7.972	1004.7E6	389.8E6	564.608	640.166
10) T	Pentachlo...	8.562	8.392	12547.7E6	4003.7E6	533.700	555.282
11) T	2,4,5-TP ...	9.106	8.753	5084.5E6	1675.7E6	556.353	581.440
12) T	2,4,5-T	9.392	9.136	4772.4E6	1498.2E6	521.833	561.353
13) T	2,4-DB	9.942	9.641	603.2E6	191.2E6	347.634	541.925 #
14) T	DINOSEB	11.087	9.968	3027.4E6	910.0E6	422.174	444.300
15) T	Picloram	10.920	10.957	5039.0E6	1773.7E6	414.822	445.121
16) T	DCPA	11.354	10.893	5586.9E6	1975.6E6	409.015	463.497

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051421\
Data File : PS015018.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 May 2021 18:08
Operator : DD\AJ
Sample : M2364-02MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
AST2-COPR4-1MSD

Manual Integrations
APPROVED

Ankita
5/17/2021 5:12:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 18:46:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051021.M
Quant Title : 8080.M
QLast Update : Tue May 11 00:04:04 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

