

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051623\
 Data File : PS023032.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2023 17:04
 Operator : AR\AJ
 Sample : PB152811BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB152811BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 21:36:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
 Quant Title : 8080.M
 QLast Update : Wed May 10 16:09:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.486	6.857	824.1E6	305.7E6	541.817	544.813
Target Compounds							
1) T	Dalapon	2.550	2.578	764.3E6	405.5E6	472.702	469.761
2) T	3,5-DICHL...	5.838	6.022	1012.5E6	429.4E6	504.547	506.248
3) T	4-Nitroph...	6.317	6.462	518.9E6	166.5E6	492.907	501.877
5) T	DICAMBA	6.635	7.024	3381.2E6	1272.1E6	514.905	510.585
6) T	MCP	6.793	7.117	222.8E6	88198302	47.681	49.849
7) T	MCPA	6.922	7.327	346.5E6	132.3E6	49.493	49.982
8) T	DICHLORPROP	7.247	7.646	963.5E6	315.2E6	514.089	509.328
9) T	2,4-D	7.449	7.934	1191.8E6	366.6E6	514.956	512.363
10) T	Pentachlo...	7.683	8.365	14390.6E6	4891.7E6	537.569	527.714
11) T	2,4,5-TP ...	8.221	8.725	6206.4E6	1852.9E6	531.319	520.448
12) T	2,4,5-T	8.483	9.099	6151.7E6	1668.2E6	532.281	519.560
13) T	2,4-DB	8.999	9.605	1193.7E6	200.7E6	508.839	522.585
14) T	DINOSEB	10.074	9.935	4496.5E6	1258.2E6	550.785	531.883
15) T	Picloram	9.895	10.901	8599.0E6	2040.4E6	501.445	479.103
16) T	DCPA	10.345	10.863	7284.3E6	1994.5E6	558.384	533.768

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051623\
Data File : PS023032.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2023 17:04
Operator : AR\AJ
Sample : PB152811BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB152811BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 21:36:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
Quant Title : 8080.M
QLast Update : Wed May 10 16:09:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

