

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051723\
 Data File : PS023040.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 May 2023 16:52
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 03:43:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
 Quant Title : 8080.M
 QLast Update : Wed May 10 16:09:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.486	6.856	1124.0E6	424.1E6	739.011	755.905
Target Compounds							
1) T	Dalapon	2.550	2.579	1100.5E6	575.9E6	680.616	667.162
2) T	3,5-DICHL...	5.837	6.022	1398.7E6	605.8E6	697.027	714.199
3) T	4-Nitroph...	6.317	6.461	729.6E6	231.2E6	693.009	696.731
5) T	DICAMBA	6.634	7.023	4808.0E6	1859.4E6	732.190	746.319
6) T	MCP	6.794	7.118	339.3E6	131.8E6	72.610	74.519
7) T	MCPA	6.923	7.328	488.8E6	182.7E6	69.818	68.984
8) T	DICHLORPROP	7.247	7.645	1309.6E6	439.7E6	698.770	710.448
9) T	2,4-D	7.449	7.933	1585.8E6	497.9E6	685.218	695.943
10) T	Pentachlo...	7.683	8.364	19187.9E6	6722.8E6	716.775	725.252
11) T	2,4,5-TP ...	8.221	8.724	8140.9E6	2515.1E6	696.934	706.447
12) T	2,4,5-T	8.482	9.098	8129.5E6	2271.5E6	703.417	707.443
13) T	2,4-DB	8.998	9.604	1702.4E6	275.4E6	725.642	717.105
14) T	DINOSEB	10.074	9.934	5862.4E6	1690.3E6	718.095	714.525
15) T	Picloram	9.894	10.901	11894.6E6	2893.5E6	693.626	679.407
16) T	DCPA	10.344	10.864	9775.9E6	2756.0E6	749.387	737.562

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS051723\
Data File : PS023040.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2023 16:52
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 03:43:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS051023.M
Quant Title : 8080.M
QLast Update : Wed May 10 16:09:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

